

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL02 Date Analyzed: 10/24/2024

Lab Sample No.: PSTDCCC050 Data File : PL092613.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.061	8.955	9.155	49.250	50.000	-1.5
Endrin	6.582	6.474	6.674	42.050	50.000	-15.9
gamma-BHC (Lindane)	4.335	4.227	4.427	46.560	50.000	-6.9
Heptachlor	4.922	4.815	5.015	45.870	50.000	-8.3
Heptachlor epoxide	5.690	5.583	5.783	43.230	50.000	-13.5
Methoxychlor	7.506	7.400	7.600	51.180	50.000	2.4
Tetrachloro-m-xylene	3.547	3.438	3.638	49.340	50.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL02 Date Analyzed: 10/24/2024

Lab Sample No.: PSTDCCC050 Data File : PL092613.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.920	7.817	8.017	53.280	50.000	6.6
Endrin	5.645	5.541	5.741	51.590	50.000	3.2
gamma-BHC (Lindane)	3.613	3.510	3.710	53.810	50.000	7.6
Heptachlor	3.952	3.849	4.049	53.550	50.000	7.1
Heptachlor epoxide	4.735	4.631	4.831	53.230	50.000	6.5
Methoxychlor	6.619	6.515	6.715	56.150	50.000	12.3
Tetrachloro-m-xylene	2.779	2.677	2.877	54.610	50.000	9.2