

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL03 Date Analyzed: 10/25/2024

Lab Sample No.: PSTDCCC050 Data File : PL092639.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.057	8.955	9.155	51.240	50.000	2.5
Endrin	6.577	6.474	6.674	43.260	50.000	-13.5
gamma-BHC (Lindane)	4.330	4.227	4.427	47.880	50.000	-4.2
Heptachlor	4.917	4.815	5.015	46.560	50.000	-6.9
Heptachlor epoxide	5.685	5.583	5.783	45.530	50.000	-8.9
Methoxychlor	7.501	7.400	7.600	51.890	50.000	3.8
Tetrachloro-m-xylene	3.541	3.438	3.638	50.750	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL03 Date Analyzed: 10/25/2024

Lab Sample No.: PSTDCCC050 Data File : PL092639.D Time Analyzed: 15:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.918	7.817	8.017	58.290	50.000	16.6
Endrin	5.643	5.541	5.741	53.930	50.000	7.9
gamma-BHC (Lindane)	3.612	3.510	3.710	55.850	50.000	11.7
Heptachlor	3.951	3.849	4.049	54.990	50.000	10.0
Heptachlor epoxide	4.734	4.631	4.831	54.980	50.000	10.0
Methoxychlor	6.617	6.515	6.715	57.220	50.000	14.4
Tetrachloro-m-xylene	2.779	2.677	2.877	56.350	50.000	12.7