

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG No.: P4460
 Instrument ID: BNA_F Calibration Date/Time: 10/24/2024 15:16
 Lab File ID: BF140001.D Init. Calib. Date(s): 10/18/2024 10/18/2024
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.278	1.208		-5.5	
Benzaldehyde	0.931	0.931		0.0	
Phenol-d6	1.655	1.521		-8.1	
Phenol	1.706	1.601		-6.2	20.0
bis(2-Chloroethyl)ether	1.319	1.259		-4.5	
2-Chlorophenol	1.306	1.259		-3.6	
2-Methylphenol	1.114	1.047		-6.0	
2,2-oxybis(1-Chloropropane)	2.248	1.948		-13.3	
Acetophenone	0.490	0.472		-3.7	
3+4-Methylphenols	1.424	1.322		-7.2	
n-Nitroso-di-n-propylamine	1.004	0.889	0.050	-11.5	
Nitrobenzene-d5	0.361	0.374		3.6	
Hexachloroethane	0.534	0.526		-1.5	
Nitrobenzene	0.396	0.397		0.3	
Isophorone	0.685	0.655		-4.4	
2-Nitrophenol	0.149	0.178		19.5	20.0
2,4-Dimethylphenol	0.249	0.238		-4.4	
bis(2-Chloroethoxy)methane	0.415	0.407		-1.9	
2,4-Dichlorophenol	0.283	0.285		0.7	20.0
Naphthalene	1.031	1.015		-1.6	
4-Chloroaniline	0.352	0.336		-4.5	
Hexachlorobutadiene	0.197	0.203		3.0	20.0
Caprolactam	0.090	0.089		-1.1	
4-Chloro-3-methylphenol	0.314	0.306		-2.5	20.0
2-Methylnaphthalene	0.632	0.623		-1.4	
Hexachlorocyclopentadiene	0.188	0.199	0.050	5.9	
2,4,6-Trichlorophenol	0.382	0.387		1.3	20.0
2-Fluorobiphenyl	1.210	1.197		-1.1	
2,4,5-Trichlorophenol	0.394	0.415		5.3	
1,1-Biphenyl	1.392	1.393		0.1	
2-Chloronaphthalene	1.121	1.127		0.5	
2-Nitroaniline	0.343	0.369		7.6	
Dimethylphthalate	1.246	1.232		-1.1	
Acenaphthylene	1.621	1.600		-1.3	
2,6-Dinitrotoluene	0.270	0.292		8.1	
3-Nitroaniline	0.278	0.285		2.5	
Acenaphthene	1.049	1.061		1.1	20.0
2,4-Dinitrophenol	0.093	0.142	0.050	52.7	
4-Nitrophenol	0.214	0.216	0.050	0.9	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG No.: P4460
 Instrument ID: BNA_F Calibration Date/Time: 10/24/2024 15:16
 Lab File ID: BF140001.D Init. Calib. Date(s): 10/18/2024 10/18/2024
 EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46
 GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.505	1.482		-1.5	
2,4-Dinitrotoluene	0.332	0.371		11.7	
Diethylphthalate	1.223	1.183		-3.3	
4-Chlorophenyl-phenylether	0.579	0.577		-0.3	
Fluorene	1.146	1.115		-2.7	
4-Nitroaniline	0.263	0.268		1.9	
4,6-Dinitro-2-methylphenol	0.082	0.115		40.2	
n-Nitrosodiphenylamine	0.599	0.597		-0.3	20.0
2,4,6-Tribromophenol	0.187	0.207		10.7	
4-Bromophenyl-phenylether	0.206	0.220		6.8	
Hexachlorobenzene	0.232	0.242		4.3	
Atrazine	0.162	0.163		0.6	
Pentachlorophenol	0.141	0.159		12.8	20.0
Phenanthrene	0.945	0.922		-2.4	
Anthracene	0.922	0.910		-1.3	
Carbazole	0.857	0.810		-5.5	
Di-n-butylphthalate	0.990	0.945		-4.5	
Fluoranthene	0.955	0.896		-6.2	20.0
Pyrene	1.751	1.855		5.9	
Terphenyl-d14	1.227	1.306		6.4	
Butylbenzylphthalate	0.525	0.544		3.6	
3,3-Dichlorobenzidine	0.380	0.435		14.5	
Benzo (a) anthracene	1.302	1.322		1.5	
Chrysene	1.194	1.125		-5.8	
Bis (2-ethylhexyl) phthalate	0.589	0.627		6.5	
Di-n-octyl phthalate	1.071	1.102		2.9	20.0
Benzo (b) fluoranthene	1.218	1.087		-10.8	
Benzo (k) fluoranthene	1.051	1.032		-1.8	
Benzo (a) pyrene	1.002	0.984		-1.8	20.0
Indeno (1,2,3-cd) pyrene	1.287	1.386		7.7	
Dibenzo (a,h) anthracene	1.073	1.154		7.5	
Benzo (g,h,i) perylene	1.072	1.167		8.9	
1,2,4,5-Tetrachlorobenzene	0.540	0.560		3.7	
1,4-Dioxane	0.596	0.573		-3.9	20.0
2,3,4,6-Tetrachlorophenol	0.309	0.321		3.9	

All other compounds must meet a minimum RRF of 0.010.