

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
 Data File : BM048211.D
 Acq On : 24 Oct 2024 11:11
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 12:35:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 18:21:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.487	0.489	-0.4	113	0.00
3	Pyridine	1.294	1.337	-3.3	117	0.00
4	n-Nitrosodimethylamine	0.549	0.569	-3.6	115	0.00
5 S	2-Fluorophenol	1.190	1.228	-3.2	117	0.00
6	Aniline	1.286	1.347	-4.7	113	0.00
7 S	Phenol-d6	1.550	1.554	-0.3	113	0.00
8	2-Chlorophenol	1.293	1.306	-1.0	115	0.00
9	Benzaldehyde	0.761	0.734	3.5	112	0.00
10 C	Phenol	1.559	1.569	-0.6	114	0.00
11	bis(2-Chloroethyl)ether	1.242	1.263	-1.7	114	0.00
12	1,3-Dichlorobenzene	1.490	1.521	-2.1	117	0.00
13 C	1,4-Dichlorobenzene	1.514	1.547	-2.2	118	0.00
14	1,2-Dichlorobenzene	1.463	1.483	-1.4	116	0.00
15	Benzyl Alcohol	0.989	0.993	-0.4	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.816	1.809	0.4	114	0.00
17	2-Methylphenol	1.036	1.046	-1.0	114	0.00
18	Hexachloroethane	0.555	0.560	-0.9	115	0.00
19 P	n-Nitroso-di-n-propylamine	0.980	0.948	3.3	109	0.00
20	3+4-Methylphenols	1.435	1.436	-0.1	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
22	Acetophenone	0.488	0.486	0.4	111	0.00
23 S	Nitrobenzene-d5	0.354	0.357	-0.8	111	0.00
24	Nitrobenzene	0.367	0.370	-0.8	112	0.00
25	Isophorone	0.649	0.635	2.2	108	0.00
26 C	2-Nitrophenol	0.169	0.177	-4.7	112	0.00
27	2,4-Dimethylphenol	0.206	0.208	-1.0	111	0.00
28	bis(2-Chloroethoxy)methane	0.411	0.416	-1.2	111	0.00
29 C	2,4-Dichlorophenol	0.296	0.311	-5.1	114	0.00
30	1,2,4-Trichlorobenzene	0.339	0.353	-4.1	117	0.00
31	Naphthalene	1.043	1.067	-2.3	115	0.00
32	Benzoic acid	0.229	0.229	0.0	105	0.00
33	4-Chloroaniline	0.338	0.365	-8.0	111	0.00
34 C	Hexachlorobutadiene	0.210	0.218	-3.8	117	0.00
35	Caprolactam	0.096	0.095	1.0	105	0.00
36 C	4-Chloro-3-methylphenol	0.309	0.312	-1.0	110	0.00
37	2-Methylnaphthalene	0.697	0.706	-1.3	113	0.00
38	1-Methylnaphthalene	0.691	0.695	-0.6	112	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	0.572	0.590	-3.1	112	0.00
41 P	Hexachlorocyclopentadiene	0.193	0.205	-6.2	112	0.00
42 S	2,4,6-Tribromophenol	0.245	0.251	-2.4	107	0.00
43 C	2,4,6-Trichlorophenol	0.379	0.399	-5.3	111	0.00
44	2,4,5-Trichlorophenol	0.423	0.437	-3.3	109	0.00
45 S	2-Fluorobiphenyl	1.303	1.338	-2.7	110	0.00
46	1,1'-Biphenyl	1.428	1.466	-2.7	111	0.00
47	2-Chloronaphthalene	1.126	1.160	-3.0	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
 Data File : BM048211.D
 Acq On : 24 Oct 2024 11:11
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 12:35:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 18:21:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.305	0.318	-4.3	106	0.00
49	Acenaphthylene	1.643	1.687	-2.7	110	0.00
50	Dimethylphthalate	1.365	1.373	-0.6	108	0.00
51	2,6-Dinitrotoluene	0.305	0.315	-3.3	108	0.00
52 C	Acenaphthene	1.044	1.059	-1.4	110	0.00
53	3-Nitroaniline	0.287	0.318	-10.8	109	0.00
54 P	2,4-Dinitrophenol	0.178	0.184	-3.4	108	0.00
55	Dibenzofuran	1.658	1.689	-1.9	109	0.00
56 P	4-Nitrophenol	0.256	0.264	-3.1	105	0.00
57	2,4-Dinitrotoluene	0.402	0.425	-5.7	107	0.00
58	Fluorene	1.311	1.335	-1.8	109	0.00
59	2,3,4,6-Tetrachlorophenol	0.351	0.360	-2.6	107	0.00
60	Diethylphthalate	1.382	1.391	-0.7	108	0.00
61	4-Chlorophenyl-phenylether	0.656	0.673	-2.6	110	0.00
62	4-Nitroaniline	0.297	0.327	-10.1	106	0.00
63	Azobenzene	1.239	1.278	-3.1	107	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	0.116	0.129	-11.2	106	0.00
66 c	n-Nitrosodiphenylamine	0.549	0.570	-3.8	107	0.00
67	4-Bromophenyl-phenylether	0.203	0.209	-3.0	107	0.00
68	Hexachlorobenzene	0.244	0.251	-2.9	107	0.00
69	Atrazine	0.154	0.154	0.0	89	0.00
70 C	Pentachlorophenol	0.160	0.171	-6.9	106	0.00
71	Phenanthrene	0.990	1.012	-2.2	106	0.00
72	Anthracene	0.959	0.994	-3.6	106	0.00
73	Carbazole	0.940	0.981	-4.4	105	0.00
74	Di-n-butylphthalate	1.170	1.227	-4.9	106	0.00
75 C	Fluoranthene	1.181	1.202	-1.8	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzidine	0.231	0.372	-61.0#	111	0.00
78	Pyrene	1.339	1.394	-4.1	105	0.00
79 S	Terphenyl-d14	0.980	1.013	-3.4	103	0.00
80	Butylbenzylphthalate	0.560	0.598	-6.8	104	0.00
81	Benzo(a)anthracene	1.278	1.318	-3.1	104	0.00
82	3,3'-Dichlorobenzidine	0.438	0.458	-4.6	99	0.00
83	Chrysene	1.227	1.249	-1.8	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.814	0.870	-6.9	103	0.00
85 c	Di-n-octyl phthalate	1.409	1.494	-6.0	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.323	1.371	-3.6	102	0.00
88	Benzo(b)fluoranthene	1.147	1.188	-3.6	103	0.00
89	Benzo(k)fluoranthene	1.109	1.158	-4.4	103	0.00
90 C	Benzo(a)pyrene	0.997	1.039	-4.2	103	0.00
91	Dibenzo(a,h)anthracene	1.101	1.147	-4.2	102	-0.01
92	Benzo(g,h,i)perylene	1.128	1.164	-3.2	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
Data File : BM048211.D
Acq On : 24 Oct 2024 11:11
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Oct 24 12:35:32 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 18:21:59 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0