

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
 Data File : BM048211.D
 Acq On : 24 Oct 2024 11:11
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 12:35:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 18:21:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	40.151	-0.4	113	0.00
3	Pyridine	40.000	41.330	-3.3	117	0.00
4	n-Nitrosodimethylamine	40.000	41.411	-3.5	115	0.00
5 S	2-Fluorophenol	80.000	82.582	-3.2	117	0.00
6	Aniline	40.000	41.906	-4.8	113	0.00
7 S	Phenol-d6	80.000	80.214	-0.3	113	0.00
8	2-Chlorophenol	40.000	40.403	-1.0	115	0.00
9	Benzaldehyde	40.000	38.579	3.6	112	0.00
10 C	Phenol	40.000	40.265	-0.7	114	0.00
11	bis(2-Chloroethyl)ether	40.000	40.673	-1.7	114	0.00
12	1,3-Dichlorobenzene	40.000	40.812	-2.0	117	0.00
13 C	1,4-Dichlorobenzene	40.000	40.870	-2.2	118	0.00
14	1,2-Dichlorobenzene	40.000	40.527	-1.3	116	0.00
15	Benzyl Alcohol	40.000	40.173	-0.4	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.836	0.4	114	0.00
17	2-Methylphenol	40.000	40.383	-1.0	114	0.00
18	Hexachloroethane	40.000	40.340	-0.9	115	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.684	3.3	109	0.00
20	3+4-Methylphenols	40.000	40.027	-0.1	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	39.843	0.4	111	0.00
23 S	Nitrobenzene-d5	80.000	80.557	-0.7	111	0.00
24	Nitrobenzene	40.000	40.350	-0.9	112	0.00
25	Isophorone	40.000	39.139	2.2	108	0.00
26 C	2-Nitrophenol	40.000	41.763	-4.4	112	0.00
27	2,4-Dimethylphenol	40.000	40.459	-1.1	111	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.472	-1.2	111	0.00
29 C	2,4-Dichlorophenol	40.000	42.073	-5.2	114	0.00
30	1,2,4-Trichlorobenzene	40.000	41.704	-4.3	117	0.00
31	Naphthalene	40.000	40.904	-2.3	115	0.00
32	Benzoic acid	40.000	39.970	0.1	105	0.00
33	4-Chloroaniline	40.000	43.162	-7.9	111	0.00
34 C	Hexachlorobutadiene	40.000	41.573	-3.9	117	0.00
35	Caprolactam	40.000	39.528	1.2	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.402	-1.0	110	0.00
37	2-Methylnaphthalene	40.000	40.534	-1.3	113	0.00
38	1-Methylnaphthalene	40.000	40.208	-0.5	112	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.260	-3.1	112	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.405	-6.0	112	0.00
42 S	2,4,6-Tribromophenol	80.000	82.055	-2.6	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.081	-5.2	111	0.00
44	2,4,5-Trichlorophenol	40.000	41.290	-3.2	109	0.00
45 S	2-Fluorobiphenyl	80.000	82.152	-2.7	110	0.00
46	1,1'-Biphenyl	40.000	41.053	-2.6	111	0.00
47	2-Chloronaphthalene	40.000	41.195	-3.0	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
 Data File : BM048211.D
 Acq On : 24 Oct 2024 11:11
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 12:35:32 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 18:21:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.638	-4.1	106	0.00
49	Acenaphthylene	40.000	41.088	-2.7	110	0.00
50	Dimethylphthalate	40.000	40.215	-0.5	108	0.00
51	2,6-Dinitrotoluene	40.000	41.379	-3.4	108	0.00
52 C	Acenaphthene	40.000	40.593	-1.5	110	0.00
53	3-Nitroaniline	40.000	44.277	-10.7	109	0.00
54 P	2,4-Dinitrophenol	40.000	41.268	-3.2	108	0.00
55	Dibenzofuran	40.000	40.753	-1.9	109	0.00
56 P	4-Nitrophenol	40.000	41.240	-3.1	105	0.00
57	2,4-Dinitrotoluene	40.000	42.253	-5.6	107	0.00
58	Fluorene	40.000	40.738	-1.8	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.036	-2.6	107	0.00
60	Diethylphthalate	40.000	40.242	-0.6	108	0.00
61	4-Chlorophenyl-phenylether	40.000	41.054	-2.6	110	0.00
62	4-Nitroaniline	40.000	43.956	-9.9	106	0.00
63	Azobenzene	40.000	41.244	-3.1	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.393	-11.0	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.515	-3.8	107	0.00
67	4-Bromophenyl-phenylether	40.000	41.136	-2.8	107	0.00
68	Hexachlorobenzene	40.000	41.163	-2.9	107	0.00
69	Atrazine	40.000	40.162	-0.4	89	0.00
70 C	Pentachlorophenol	40.000	42.774	-6.9	106	0.00
71	Phenanthrene	40.000	40.859	-2.1	106	0.00
72	Anthracene	40.000	41.464	-3.7	106	0.00
73	Carbazole	40.000	41.754	-4.4	105	0.00
74	Di-n-butylphthalate	40.000	41.970	-4.9	106	0.00
75 C	Fluoranthene	40.000	40.698	-1.7	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzidine	40.000	64.389	-61.0#	111	0.00
78	Pyrene	40.000	41.655	-4.1	105	0.00
79 S	Terphenyl-d14	80.000	82.723	-3.4	103	0.00
80	Butylbenzylphthalate	40.000	42.743	-6.9	104	0.00
81	Benzo(a)anthracene	40.000	41.256	-3.1	104	0.00
82	3,3'-Dichlorobenzidine	40.000	41.868	-4.7	99	0.00
83	Chrysene	40.000	40.739	-1.8	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.782	-7.0	103	0.00
85 c	Di-n-octyl phthalate	40.000	42.418	-6.0	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.467	-3.7	102	0.00
88	Benzo(b)fluoranthene	40.000	41.448	-3.6	103	0.00
89	Benzo(k)fluoranthene	40.000	41.789	-4.5	103	0.00
90 C	Benzo(a)pyrene	40.000	41.695	-4.2	103	0.00
91	Dibenzo(a,h)anthracene	40.000	41.662	-4.2	102	-0.01
92	Benzo(g,h,i)perylene	40.000	41.271	-3.2	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
Data File : BM048211.D
Acq On : 24 Oct 2024 11:11
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Oct 24 12:35:32 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 18:21:59 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0