

Surrogate Summary

SW-846

SDG No.: P4460

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4460-02	WB-303-TOP	2-Fluorophenol	150	84.0	56		30 (18)	130 (112)
		Phenol-d6	150	82.9	55		30 (15)	130 (107)
		Nitrobenzene-d5	100	59.6	60		30 (18)	130 (107)
		2-Fluorobiphenyl	100	55.1	55		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	82.2	55		30 (10)	130 (116)
		Terphenyl-d14	100	54.1	54		30 (10)	130 (105)
P4460-03	WB-303-BOT	2-Fluorophenol	150	127	85		30 (18)	130 (112)
		Phenol-d6	150	126	84		30 (15)	130 (107)
		Nitrobenzene-d5	100	101	101		30 (18)	130 (107)
		2-Fluorobiphenyl	100	93.6	94		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	143	95		30 (10)	130 (116)
		Terphenyl-d14	100	101	101		30 (10)	130 (105)
P4460-03MS	WB-303-BOTMS	2-Fluorophenol	150	119	79		30 (18)	130 (112)
		Phenol-d6	150	119	79		30 (15)	130 (107)
		Nitrobenzene-d5	100	92.8	93		30 (18)	130 (107)
		2-Fluorobiphenyl	100	90.6	91		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	140	94		30 (10)	130 (116)
		Terphenyl-d14	100	77.7	78		30 (10)	130 (105)
P4460-03MSD	WB-303-BOTMSD	2-Fluorophenol	150	113	75		30 (18)	130 (112)
		Phenol-d6	150	140	93		30 (15)	130 (107)
		Nitrobenzene-d5	100	101	101		30 (18)	130 (107)
		2-Fluorobiphenyl	100	98.3	98		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	149	100		30 (10)	130 (116)
		Terphenyl-d14	100	81.9	82		30 (10)	130 (105)
PB164286BL	PB164286BL	2-Fluorophenol	150	134	90		30 (18)	130 (112)
		Phenol-d6	150	131	87		30 (15)	130 (107)
		Nitrobenzene-d5	100	99.4	99		30 (18)	130 (107)
		2-Fluorobiphenyl	100	94.7	95		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	156	104		30 (10)	130 (116)
		Terphenyl-d14	100	98.4	98		30 (10)	130 (105)
PB164286BS	PB164286BS	2-Fluorophenol	150	123	82		30 (18)	130 (112)
		Phenol-d6	150	120	80		30 (15)	130 (107)
		Nitrobenzene-d5	100	91.8	92		30 (18)	130 (107)
		2-Fluorobiphenyl	100	86.5	87		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	147	98		30 (10)	130 (116)
		Terphenyl-d14	100	95.7	96		30 (10)	130 (105)

Surrogate Summary

SW-846

SDG No.: P4460

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4460-06	WB-303-SW	2-Fluorophenol	75	42.1	56		15 (10)	110 (139)
		Phenol-d6	75	36.5	49		15 (10)	110 (134)
		Nitrobenzene-d5	50	42.7	85		30 (49)	130 (133)
		2-Fluorobiphenyl	50	40.3	81		30 (52)	130 (132)
		2,4,6-Tribromophenol	75	67.4	90		15 (44)	110 (137)
		Terphenyl-d14	50	45.8	92		30 (48)	130 (125)
PB164369BL	PB164369BL	2-Fluorophenol	150	151	101		15 (10)	110 (139)
		Phenol-d6	150	138	92		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.1	92		30 (49)	130 (133)
		2-Fluorobiphenyl	100	93.8	94		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	88		15 (44)	110 (137)
		Terphenyl-d14	100	105	105		30 (48)	130 (125)
PB164369BS	PB164369BS	2-Fluorophenol	150	144	96		15 (10)	110 (139)
		Phenol-d6	150	131	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	96.5	96		30 (49)	130 (133)
		2-Fluorobiphenyl	100	99.0	99		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	89		15 (44)	110 (137)
		Terphenyl-d14	100	103	103		30 (48)	130 (125)
PB164369BSD	PB164369BSD	2-Fluorophenol	150	146	97		15 (10)	110 (139)
		Phenol-d6	150	135	90		15 (10)	110 (134)
		Nitrobenzene-d5	100	94.8	95		30 (49)	130 (133)
		2-Fluorobiphenyl	100	96.5	96		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	132	88		15 (44)	110 (137)
		Terphenyl-d14	100	110	110		30 (48)	130 (125)