

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4460

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BM048233.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		
							Qual	Qual	Low	High	RPD
PB164369BSD	Benzaldehyde	50	14.2	ug/L	28	4			20 (10)	160 (162)	20 (20)
	Phenol	50	49.6	ug/L	99	4			20 (66)	160 (118)	20 (20)
	bis(2-Chloroethyl)ether	50	48.8	ug/L	98	1			70 (62)	130 (103)	20 (20)
	2-Chlorophenol	50	50.0	ug/L	100	3			70 (70)	130 (117)	20 (20)
	2-Methylphenol	50	48.8	ug/L	98	5			70 (69)	130 (109)	20 (20)
	2,2-oxybis(1-Chloropropane)	50	49.4	ug/L	99	0			70 (65)	130 (100)	20 (20)
	Acetophenone	50	47.9	ug/L	96	1			70 (60)	130 (104)	20 (20)
	3+4-Methylphenols	50	47.4	ug/L	95	5			20 (67)	160 (106)	20 (20)
	N-Nitroso-di-n-propylamine	50	44.8	ug/L	90	2			70 (57)	130 (107)	20 (20)
	Hexachloroethane	50	46.1	ug/L	92	2			20 (76)	160 (118)	20 (20)
	Nitrobenzene	50	46.8	ug/L	94	2			70 (58)	130 (106)	20 (20)
	Isophorone	50	47.3	ug/L	95	1			70 (61)	130 (102)	20 (20)
	2-Nitrophenol	50	50.6	ug/L	101	3			70 (70)	130 (115)	20 (20)
	2,4-Dimethylphenol	50	60.9	ug/L	122	2			70 (42)	130 (142)	20 (20)
	bis(2-Chloroethoxy)methane	50	48.6	ug/L	97	1			70 (58)	130 (109)	20 (20)
	2,4-Dichlorophenol	50	50.0	ug/L	100	2			70 (66)	130 (115)	20 (20)
	Naphthalene	50	46.7	ug/L	93	1			70 (64)	130 (107)	20 (20)
	4-Chloroaniline	50	20.7	ug/L	41	4	*		70 (10)	130 (85)	20 (20)
	Hexachlorobutadiene	50	46.5	ug/L	93	2			70 (69)	130 (101)	20 (20)
	Caprolactam	50	41.5	ug/L	83	5			20 (58)	160 (128)	20 (20)
	4-Chloro-3-methylphenol	50	45.9	ug/L	92	4			70 (65)	130 (114)	20 (20)
	2-Methylnaphthalene	50	46.5	ug/L	93	1			70 (64)	130 (107)	20 (20)
	Hexachlorocyclopentadiene	100	210	ug/L	210	0	*		20 (36)	160 (160)	20 (20)
	2,4,6-Trichlorophenol	50	51.8	ug/L	104	1			70 (61)	130 (110)	20 (20)
	2,4,5-Trichlorophenol	50	49.7	ug/L	99	1			70 (70)	130 (106)	20 (20)
	1,1-Biphenyl	50	49.3	ug/L	99	2			70 (72)	130 (98)	20 (20)
	2-Chloronaphthalene	50	48.5	ug/L	97	2			70 (59)	130 (106)	20 (20)
	2-Nitroaniline	50	50.2	ug/L	100	1			70 (73)	130 (114)	20 (20)
	Dimethylphthalate	50	47.4	ug/L	95	0			70 (64)	130 (103)	20 (20)
	Acenaphthylene	50	51.7	ug/L	103	1			70 (79)	130 (103)	20 (20)
	2,6-Dinitrotoluene	50	46.0	ug/L	92	1			70 (64)	130 (110)	20 (20)
	3-Nitroaniline	50	29.2	ug/L	58	2	*		70 (28)	130 (100)	20 (20)
	Acenaphthene	50	48.8	ug/L	98	2			70 (59)	130 (113)	20 (20)
	2,4-Dinitrophenol	100	96.2	ug/L	96	3			20 (36)	160 (166)	20 (20)
	4-Nitrophenol	100	96.1	ug/L	96	1			20 (45)	160 (147)	20 (20)
	Dibenzofuran	50	47.5	ug/L	95	2			70 (65)	130 (106)	20 (20)
	2,4-Dinitrotoluene	50	46.6	ug/L	93	2			70 (60)	130 (115)	20 (20)
	Diethylphthalate	50	46.1	ug/L	92	2			70 (63)	130 (105)	20 (20)
	4-Chlorophenyl-phenylether	50	46.7	ug/L	93	1			70 (61)	130 (104)	20 (20)
	Fluorene	50	45.8	ug/L	92	3			70 (64)	130 (107)	20 (20)
	4-Nitroaniline	50	45.4	ug/L	91	1			70 (55)	130 (125)	20 (20)
	4,6-Dinitro-2-methylphenol	50	52.3	ug/L	105	4			70 (62)	130 (132)	20 (20)
	N-Nitrosodiphenylamine	50	52.9	ug/L	106	0			70 (61)	130 (109)	20 (20)
	4-Bromophenyl-phenylether	50	50.7	ug/L	101	1			70 (73)	130 (103)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4460

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BM048233.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		
							Qual	Qual	Low	High	RPD
PB164369BSD	Hexachlorobenzene	50	49.4	ug/L	99	0			70 (73)	130 (106)	20 (20)
	Atrazine	50	64.2	ug/L	128	1			70 (76)	130 (120)	20 (20)
	Pentachlorophenol	100	97.2	ug/L	97	1			20 (47)	160 (114)	20 (20)
	Phenanthrene	50	50.2	ug/L	100	1			70 (62)	130 (109)	20 (20)
	Anthracene	50	52.6	ug/L	105	2			70 (65)	130 (110)	20 (20)
	Carbazole	50	47.5	ug/L	95	3			70 (62)	130 (106)	20 (20)
	Di-n-butylphthalate	50	48.6	ug/L	97	1			70 (64)	130 (106)	20 (20)
	Fluoranthene	50	44.3	ug/L	89	5			70 (64)	130 (110)	20 (20)
	Pyrene	50	54.5	ug/L	109	8			70 (71)	130 (103)	20 (20)
	Butylbenzylphthalate	50	52.8	ug/L	106	6			70 (61)	130 (105)	20 (20)
	3,3-Dichlorobenzidine	50	34.7	ug/L	69	7	*		70 (43)	130 (108)	20 (20)
	Benzo(a)anthracene	50	50.9	ug/L	102	1			70 (62)	130 (107)	20 (20)
	Chrysene	50	50.1	ug/L	100	0			70 (61)	130 (108)	20 (20)
	bis(2-Ethylhexyl)phthalate	50	51.9	ug/L	104	3			70 (59)	130 (110)	20 (20)
	Di-n-octyl phthalate	50	49.5	ug/L	99	0			70 (52)	130 (139)	20 (20)
	Benzo(b)fluoranthene	50	51.5	ug/L	103	4			70 (77)	130 (113)	20 (20)
	Benzo(k)fluoranthene	50	52.7	ug/L	105	2			70 (77)	130 (105)	20 (20)
	Benzo(a)pyrene	50	55.5	ug/L	111	1			70 (72)	130 (131)	20 (20)
	Indeno(1,2,3-cd)pyrene	50	54.1	ug/L	108	3			70 (72)	130 (105)	20 (20)
	Dibenz(a,h)anthracene	50	53.6	ug/L	107	3			70 (78)	130 (115)	20 (20)
	Benzo(g,h,i)perylene	50	48.9	ug/L	98	3			70 (75)	130 (118)	20 (20)
	1,2,4,5-Tetrachlorobenzene	50	50.6	ug/L	101	3			70 (72)	130 (101)	20 (20)
	1,4-Dioxane	50	40.0	ug/L	80	8			20 (38)	160 (125)	20 (20)
	2,3,4,6-Tetrachlorophenol	50	48.3	ug/L	97	0			70 (63)	130 (116)	20 (20)