

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/23/2024

Lab File ID: BF139952.D Time Analyzed: 09:20

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	156634	6.893	590130	8.18	329153	9.93
UPPER LIMIT	313268	7.393	1180260	8.675	658306	10.428
LOWER LIMIT	78317	6.393	295065	7.675	164577	9.428
EPA SAMPLE NO.						
01 PB164286BL	146341	6.89	564292	8.17	320476	9.93
02 PB164286BS	158593	6.89	620873	8.18	350222	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.