

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/24/2024

Lab File ID: BF140001.D Time Analyzed: 15:16

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	162461	6.892	593909	8.18	325446	9.93
UPPER LIMIT	324922	7.392	1187820	8.675	650892	10.428
LOWER LIMIT	81230.5	6.392	296955	7.675	162723	9.428
EPA SAMPLE NO.						
01 WB-303-BOTMS	132812	6.89	485291	8.18	253613	9.93
02 WB-303-BOTMSD	114465	6.89	337350	8.18	176753	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.