

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/26/2024

Lab File ID: BF140050.D Time Analyzed: 10:38

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	128827	6.892	465620	8.18	252523	9.93
UPPER LIMIT	257654	7.392	931240	8.675	505046	10.428
LOWER LIMIT	64413.5	6.392	232810	7.675	126262	9.428
EPA SAMPLE NO.						
01 WB-303-TOP	142022	6.89	554829	8.17	315069	9.92
02 WB-303-BOT	129197	6.89	476862	8.17	260201	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.