

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/24/2024

Lab File ID: BM048211.D Time Analyzed: 11:11

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	180168	7.722	716977	10.51	442164	14.37
UPPER LIMIT	360336	8.222	1433950	11.01	884328	14.868
LOWER LIMIT	90084	7.222	358489	10.01	221082	13.868
EPA SAMPLE NO.						
01 PB164369BL	201808	7.72	755271	10.51	451085	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.