

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/25/2024

Lab File ID: BM048228.D Time Analyzed: 08:49

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	925065	17.104	823670	21.333	921417	24.28
UPPER LIMIT	1850130	17.604	1647340	21.833	1842830	24.78
LOWER LIMIT	462533	16.604	411835	20.833	460709	23.78
EPA SAMPLE NO.						
01 WB-303-SW	1247450	17.10	1150080	21.33	1229810	24.28
02 PB164369BSD	916015	17.10	699111	21.33	737917	24.28
03 PB164369BS	712540	17.10	616038	21.33	708876	24.28

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.