

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4460 SDG NO.: P4460

Lab File ID: BM048227.D

DFTPP Injection Date: 10/25/2024

Instrument ID: BNA_M

DFTPP Injection Time: 08:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.1
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	34.9
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	44.7
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.9
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	13.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.3 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM048228.D	10/25/2024	08:49
PB164369BSD	PB164369BSD	BM048233.D	10/25/2024	12:58
PB164369BS	PB164369BS	BM048237.D	10/25/2024	15:42
WB-303-SW	P4460-06	BM048238.D	10/25/2024	16:21