

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG No.: P4460

Instrument ID: MSVOA_N Calibration Date/Time: 10/22/2024 12:50

Lab File ID: VN084450.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.654	0.602		-7.95	20
1,1-Dichloroethene	0.563	0.575		2.13	20
2-Butanone	0.434	0.422		-2.77	20
Carbon Tetrachloride	0.525	0.551		4.95	20
Chloroform	1.175	1.210		2.98	20
Benzene	1.492	1.558		4.42	20
1,2-Dichloroethane	0.506	0.521		2.96	20
Trichloroethene	0.348	0.369		6.03	20
Tetrachloroethene	0.348	0.359		3.16	20
Chlorobenzene	1.109	1.161	0.3	4.69	20
1,2-Dichloroethane-d4	0.741	0.701		-5.4	20
Dibromofluoromethane	0.330	0.335		1.51	20
Toluene-d8	1.213	1.224		0.91	20
4-Bromofluorobenzene	0.442	0.456		3.17	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.