



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06  
Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG No.: P4460  
Instrument ID: \_\_\_\_\_ Calibration Date/Time: \_\_\_\_\_  
Lab File ID: \_\_\_\_\_ Init. Calib. Date(s): \_\_\_\_\_  
Heated Purge: (Y/N) \_\_\_\_\_ Init. Calib. Time(s): \_\_\_\_\_  
GC Column: \_\_\_\_\_ ID: \_\_\_\_\_ (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
				20

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.