

Surrogate Summary

SDG No.: P4460

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4460-04	WB-303-BOT	1,2-Dichloroethane-d4	50	49.6	99	70 (74)	130 (125)
		Dibromofluoromethane	50	49.6	99	70 (75)	130 (124)
		Toluene-d8	50	50.5	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.5	91	70 (77)	130 (121)

Surrogate Summary

SDG No.: **P4460**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN1022WBL01	VN1022WBL01	1,2-Dichloroethane-d4	50	45.7	91	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	49.3	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.0	90	70 (77)	130 (121)
VN1022WBS01	VN1022WBS01	1,2-Dichloroethane-d4	50	46.2	92	70 (74)	130 (125)
		Dibromofluoromethane	50	49.5	99	70 (75)	130 (124)
		Toluene-d8	50	48.8	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.0	98	70 (77)	130 (121)
VN1022WBSD0	VN1022WBSD01	1,2-Dichloroethane-d4	50	49.1	98	70 (74)	130 (125)
		Dibromofluoromethane	50	49.5	99	70 (75)	130 (124)
		Toluene-d8	50	48.0	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)

