

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** P4460

**Client:** Portal Partners Tri-Venture

**Analytical Method:** SW8260-Low

**Datafile :** VN084454.D

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low     | Limits<br>High | RPD     |
|---------------|----------------------|-------|--------|------|-----|-----|------|---------|----------------|---------|
| VN1022WBSD01  | Vinyl chloride       | 20    | 16.5   | ug/L | 83  | 0   |      | 70 (65) | 130 (117)      | 20 (20) |
|               | 1,1-Dichloroethene   | 20    | 17.9   | ug/L | 90  | 1   |      | 70 (74) | 130 (110)      | 20 (20) |
|               | 2-Butanone           | 100   | 92.3   | ug/L | 92  | 4   |      | 40 (65) | 160 (122)      | 20 (20) |
|               | Carbon Tetrachloride | 20    | 18.6   | ug/L | 93  | 1   |      | 70 (77) | 130 (113)      | 20 (20) |
|               | Chloroform           | 20    | 19.4   | ug/L | 97  | 3   |      | 70 (79) | 130 (113)      | 20 (20) |
|               | Benzene              | 20    | 18.9   | ug/L | 95  | 1   |      | 70 (82) | 130 (109)      | 20 (20) |
|               | 1,2-Dichloroethane   | 20    | 19.6   | ug/L | 98  | 6   |      | 70 (80) | 130 (115)      | 20 (20) |
|               | Trichloroethene      | 20    | 18.8   | ug/L | 94  | 1   |      | 70 (77) | 130 (113)      | 20 (20) |
|               | Tetrachloroethene    | 20    | 19.2   | ug/L | 96  | 4   |      | 70 (67) | 130 (123)      | 20 (20) |
|               | Chlorobenzene        | 20    | 19.6   | ug/L | 98  | 4   |      | 70 (82) | 130 (109)      | 20 (20) |