

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460  
 Lab File ID: VN084450.D Date Analyzed: 10/22/2024  
 Instrument ID: MSVOA\_N Time Analyzed: 12:50  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|------|---------------|--------|
| 12 HOUR STD    | 171974        | 8.22  | 293014        | 9.10 | 264421        | 11.87  |
| UPPER LIMIT    | 343948        | 8.724 | 586028        | 9.6  | 528842        | 12.365 |
| LOWER LIMIT    | 85987         | 7.724 | 146507        | 8.6  | 132211        | 11.365 |
| EPA SAMPLE NO. |               |       |               |      |               |        |
| WB-303-BOT     | 163501        | 8.22  | 298176        | 9.10 | 267301        | 11.87  |
| VN1022WBL01    | 195991        | 8.22  | 325943        | 9.10 | 276681        | 11.87  |
| VN1022WBS01    | 177710        | 8.22  | 304170        | 9.10 | 270028        | 11.87  |
| VN1022WBSD01   | 162648        | 8.22  | 284290        | 9.10 | 251894        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.