

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460
 Lab File ID: VN084450.D Date Analyzed: 10/22/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	131065	13.788				
UPPER LIMIT	262130	14.288				
LOWER LIMIT	65532.5	13.288				
EPA SAMPLE NO.						
WB-303-BOT	97626	13.79				
VN1022WBL01	106919	13.79				
VN1022WBS01	132421	13.79				
VN1022WBSD01	120786	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.