

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB164378BS

Contract: TETR16

Lab Code: CHEM **Case No.:** P4462

SAS No.: P4462

SDG NO.: P4462

Lab Sample ID: PB164378BS

Date(s) Analyzed: 10/24/2024 10/24/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP **ID:** 0.32 (mm)

GC Column:(2): RTX-CLP2 **ID:** 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-TP(Silvex)	1	9.03	8.98	9.08	4.90	9.7
	2	9.72	9.67	9.77	5.40	
2,4-D	1	8.18	8.13	8.23	4.90	2
	2	8.84	8.79	8.89	5.00	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WB-301-BOTMS

Contract: TETR16

Lab Code: CHEM Case No.: P4462

SAS No.: P4462 SDG NO.: P4462

Lab Sample ID: P4397-06MS

Date(s) Analyzed: 10/24/2024 10/24/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-TP(Silvex)	1	9.03	8.98	9.08	47.3	76.6
	2	9.73	9.68	9.78	106	
2,4-D	1	8.18	8.13	8.23	49.0	11.9
	2	8.84	8.79	8.89	55.2	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WB-301-BOTMSD

Contract: TETR16

Lab Code: CHEM Case No.: P4462

SAS No.: P4462 SDG NO.: P4462

Lab Sample ID: P4397-06MSD

Date(s) Analyzed: 10/24/2024 10/24/2024

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.18	8.13	8.23	49.2	14
	2	8.83	8.78	8.88	56.6	
2,4,5-TP(Silvex)	1	9.03	8.98	9.08	47.5	81.6
	2	9.73	9.68	9.78	113	