

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4462 SAS No.: P4462 SDG NO.: P4462

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL01 Date Analyzed: 10/24/2024

Lab Sample No.: HSTDCCC750 Data File : PS028036.D Time Analyzed: 11:25

| COMPOUND         | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|-------|-----------|-------|--------------------|-------------------|------|
|                  |       | FROM      | TO    |                    |                   |      |
| 2,4,5-TP(Silvex) | 9.028 | 8.929     | 9.129 | 684.860            | 712.500           | -3.9 |
| 2,4-D            | 8.176 | 8.077     | 8.277 | 665.580            | 705.000           | -5.6 |
| 2,4-DCAA         | 7.091 | 6.991     | 7.191 | 709.960            | 750.000           | -5.3 |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4462 SAS No.: P4462 SDG NO.: P4462

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL01 Date Analyzed: 10/24/2024

Lab Sample No.: HSTDCCC750 Data File : PS028036.D Time Analyzed: 11:25

| COMPOUND         | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|-------|----------------------|-------|--------------------|-------------------|------|
| 2,4,5-TP(Silvex) | 9.725 | 9.625                | 9.825 | 736.810            | 712.500           | 3.4  |
| 2,4-D            | 8.835 | 8.737                | 8.937 | 709.890            | 705.000           | 0.7  |
| 2,4-DCAA         | 7.613 | 7.514                | 7.714 | 733.180            | 750.000           | -2.2 |