

Date Collected:

Date Received:

SDG No.:

Matrix:

% Solid:

Final Vol:

Date Analyzed

Injection Volume:

Test:

10/24/24

10/24/24

P4462

TCLP

10000

TCLP Herbicide

Decanted:

иL

Prep Batch ID



Report of Analysis

Client: Tetra Tech, EMI

Project: R36724

Client Sample ID: PIBLK-PS028046.D

Lab Sample ID: I.BLK-PS028046.D

Analytical Method: SW8151A

Sample Wt/Vol: 1000 Units: mL

Soil Aliquot Vol: uL

Extraction Type:

PS028046.D

GPC Factor: 1.0 PH:

Prep Method: SW3510C

File ID/Qc Batch: Dilution:

1 10/24/24 PS102424

Qualifier MDL LOQ / CRQL Units **CAS Number Parameter** Conc. **TARGETS** 94-75-7 2,4-D 0.49 U 0.49 2.00 ug/L 93-72-1 2,4,5-TP (Silvex) 0.45 U 0.45 2.00 ug/L **SURROGATES** 19719-28-9 2,4-DCAA 506 39 - 175 101% SPK: 500

Prep Date

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit