

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102324

Review By	Abdul	Review On	10/24/2024 2:22:53 PM
Supervise By	Ankita	Supervise On	10/24/2024 2:25:32 PM
SubDirectory	PS102324	HP Acquire Method	HP Processing Method ps102324 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028006.D	23 Oct 2024 10:40		AR\AJ	Ok
2	I.BLK	I.BLK	PS028007.D	23 Oct 2024 11:04		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS028008.D	23 Oct 2024 11:28		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS028009.D	23 Oct 2024 11:52		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS028010.D	23 Oct 2024 12:16		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS028011.D	23 Oct 2024 12:40		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS028012.D	23 Oct 2024 13:04		AR\AJ	Ok
8	HSTDICV750	ICVPS102324	PS028013.D	23 Oct 2024 13:28		AR\AJ	Ok
9	I.BLK	I.BLK	PS028014.D	23 Oct 2024 14:14		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS028015.D	23 Oct 2024 14:38		AR\AJ	Ok,M
11	P4443-01	OG-315-HR-502-COMF	PS028016.D	23 Oct 2024 15:02		AR\AJ	Ok,M
12	P4443-06	OG-315-HR-502-COMF	PS028017.D	23 Oct 2024 15:26		AR\AJ	Ok,M
13	P4458-01	280517	PS028018.D	23 Oct 2024 15:49		AR\AJ	Ok,M
14	P4458-01MS	280517MS	PS028019.D	23 Oct 2024 16:13	Some compound recovery fail	AR\AJ	Ok,M
15	P4458-01MSD	280517MSD	PS028020.D	23 Oct 2024 16:37	Some compound recovery fail , RPD is high in MCPA	AR\AJ	Ok,M
16	PB174307BL	PB174307BL	PS028021.D	23 Oct 2024 17:01	Typo PB164307BL	AR\AJ	Ok,M
17	PB174307BS	PB174307BS	PS028022.D	23 Oct 2024 17:25	Typo PB164307BS	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102324

Review By	Abdul	Review On	10/24/2024 2:22:53 PM
Supervise By	Ankita	Supervise On	10/24/2024 2:25:32 PM
SubDirectory	PS102324	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	I.BLK	I.BLK	PS028023.D	23 Oct 2024 18:14		AR\AJ	Ok
19	HSTDCCC750	HSTDCCC750	PS028024.D	23 Oct 2024 18:38		AR\AJ	Ok,M
20	P4468-03	ETGI-329	PS028025.D	23 Oct 2024 19:02		AR\AJ	Ok,M
21	P4468-05	ETGI-345	PS028026.D	23 Oct 2024 19:26		AR\AJ	Ok
22	P4467-01	TP-1	PS028027.D	23 Oct 2024 19:50		AR\AJ	Ok
23	I.BLK	I.BLK	PS028028.D	24 Oct 2024 01:27		AR\AJ	Ok
24	HSTDCCC750	HSTDCCC750	PS028029.D	24 Oct 2024 01:51		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102424

Review By	Abdul	Review On	10/25/2024 3:28:01 PM
Supervise By	Ankita	Supervise On	10/28/2024 9:12:45 AM
SubDirectory	PS102424	HP Acquire Method	HP Processing Method ps102324 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC Internal Standard/PEM	PP23462
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028030.D	24 Oct 2024 09:02		AR\AJ	Ok
2	I.BLK	I.BLK	PS028031.D	24 Oct 2024 09:26		AR\AJ	Ok,M
3	HSTDCCC750	HSTDCCC750	PS028032.D	24 Oct 2024 09:50		AR\AJ	Ok
4	P4472-01	BP-F-28	PS028033.D	24 Oct 2024 10:13		AR\AJ	Ok
5	P4472-05	BP-F-6	PS028034.D	24 Oct 2024 10:37		AR\AJ	Ok
6	I.BLK	I.BLK	PS028035.D	24 Oct 2024 11:01		AR\AJ	Ok,M
7	HSTDCCC750	HSTDCCC750	PS028036.D	24 Oct 2024 11:25		AR\AJ	Ok
8	PB164378BL	PB164378BL	PS028037.D	24 Oct 2024 17:09		AR\AJ	Ok,M
9	PB164378BS	PB164378BS	PS028038.D	24 Oct 2024 17:33		AR\AJ	Ok
10	PB164261TB	PB164261TB	PS028039.D	24 Oct 2024 17:57	surrogate fail in 2nd column	AR\AJ	Not Ok
11	PB164336TB	PB164336TB	PS028040.D	24 Oct 2024 18:21	surrogate fail in 2nd column	AR\AJ	Not Ok
12	P4397-06	WB-301-BOT	PS028041.D	24 Oct 2024 18:45	both surrogate fail confirms with ms/msd	AR\AJ	Ok,M
13	P4397-06MS	WB-301-BOTMS	PS028042.D	24 Oct 2024 19:09	both surrogate fail , some comp recovery fails	AR\AJ	Ok,M
14	P4397-06MSD	WB-301-BOTMSD	PS028043.D	24 Oct 2024 19:32	both surrogate fail , some comp recovery fails	AR\AJ	Ok,M
15	P4460-04	WB-303-BOT	PS028044.D	24 Oct 2024 19:56	both Surrogate Fail	AR\AJ	Not Ok
16	P4462-02	C0AL2	PS028045.D	24 Oct 2024 20:20		AR\AJ	Ok
17	I.BLK	I.BLK	PS028046.D	24 Oct 2024 20:44		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102424

Review By	Abdul	Review On	10/25/2024 3:28:01 PM
Supervise By	Ankita	Supervise On	10/28/2024 9:12:45 AM
SubDirectory	PS102424	HP Acquire Method	HP Processing Method ps102324 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC Internal Standard/PEM	PP23462
ICV/I.BLK Surrogate Standard	PP23469
MS/MSD Standard	
LCS Standard	

18	HSTDCCC750	HSTDCCC750	PS028047.D	24 Oct 2024 21:08		ARIAJ	Ok
----	------------	------------	------------	-------------------	--	-------	----

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102824

Review By	Abdul	Review On	10/29/2024 8:52:20 AM
Supervise By	Ankita	Supervise On	10/29/2024 8:52:48 AM
SubDirectory	PS102824	HP Acquire Method	HP Processing Method ps102324 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028070.D	28 Oct 2024 09:44		AR\AJ	Ok
2	I.BLK	I.BLK	PS028071.D	28 Oct 2024 10:08		AR\AJ	Ok,M
3	HSTDCCC750	HSTDCCC750	PS028072.D	28 Oct 2024 10:32		AR\AJ	Ok,M
4	PB164261TB	PB164261TB	PS028073.D	28 Oct 2024 13:11		AR\AJ	Ok,M
5	PB164336TB	PB164336TB	PS028074.D	28 Oct 2024 13:35		AR\AJ	Ok,M
6	I.BLK	I.BLK	PS028075.D	28 Oct 2024 13:59		AR\AJ	Ok
7	HSTDCCC750	HSTDCCC750	PS028076.D	28 Oct 2024 17:17		AR\AJ	Ok

M : Manual Integration