

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	526	510	103	95 - 105	P	10/22/2024	15:01	LB133061
	Iron	10200	10000	102	95 - 105	P	10/22/2024	15:01	LB133061
	Lead	979	1000	98	95 - 105	P	10/22/2024	15:01	LB133061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.9	20.0	110	80 - 120	P	10/22/2024	15:16	LB133061
	Iron	103	100	103	80 - 120	P	10/22/2024	15:16	LB133061
	Lead	12.4	12.0	104	80 - 120	P	10/22/2024	15:16	LB133061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P4464

Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	10/22/2024	15:52	LB133061
	Iron	5010	5000	100	90 - 110	P	10/22/2024	15:52	LB133061
	Lead	5020	5000	100	90 - 110	P	10/22/2024	15:52	LB133061
CCV02	Copper	1250	1250	100	90 - 110	P	10/22/2024	16:30	LB133061
	Iron	4790	5000	96	90 - 110	P	10/22/2024	16:30	LB133061
	Lead	5030	5000	100	90 - 110	P	10/22/2024	16:30	LB133061
CCV03	Copper	1270	1250	101	90 - 110	P	10/22/2024	17:25	LB133061
	Iron	5020	5000	100	90 - 110	P	10/22/2024	17:25	LB133061
	Lead	5050	5000	101	90 - 110	P	10/22/2024	17:25	LB133061
CCV04	Copper	1260	1250	101	90 - 110	P	10/22/2024	18:19	LB133061
	Iron	5120	5000	102	90 - 110	P	10/22/2024	18:19	LB133061
	Lead	5000	5000	100	90 - 110	P	10/22/2024	18:19	LB133061
CCV05	Copper	1260	1250	101	90 - 110	P	10/22/2024	19:13	LB133061
	Iron	4880	5000	98	90 - 110	P	10/22/2024	19:13	LB133061
	Lead	5020	5000	100	90 - 110	P	10/22/2024	19:13	LB133061
CCV06	Copper	1270	1250	102	90 - 110	P	10/22/2024	20:24	LB133061
	Iron	4870	5000	97	90 - 110	P	10/22/2024	20:24	LB133061
	Lead	5030	5000	101	90 - 110	P	10/22/2024	20:24	LB133061
CCV07	Copper	1260	1250	101	90 - 110	P	10/22/2024	21:16	LB133061
	Iron	5050	5000	101	90 - 110	P	10/22/2024	21:16	LB133061
	Lead	5000	5000	100	90 - 110	P	10/22/2024	21:16	LB133061
CCV08	Copper	1260	1250	101	90 - 110	P	10/22/2024	22:29	LB133061
	Iron	4760	5000	95	90 - 110	P	10/22/2024	22:29	LB133061
	Lead	4930	5000	99	90 - 110	P	10/22/2024	22:29	LB133061
CCV09	Copper	1270	1250	102	90 - 110	P	10/22/2024	23:20	LB133061
	Iron	4850	5000	97	90 - 110	P	10/22/2024	23:20	LB133061
	Lead	5050	5000	101	90 - 110	P	10/22/2024	23:20	LB133061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	513	510	101	95 - 105	P	10/23/2024	14:59	LB133086
	Iron	10400	10000	104	95 - 105	P	10/23/2024	14:59	LB133086
	Lead	1010	1000	101	95 - 105	P	10/23/2024	14:59	LB133086

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.6	20.0	108	80 - 120	P	10/23/2024	15:11	LB133086
	Iron	104	100	104	80 - 120	P	10/23/2024	15:11	LB133086
	Lead	11.9	12.0	99	80 - 120	P	10/23/2024	15:11	LB133086

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	10/23/2024	15:57	LB133086
	Iron	4910	5000	98	90 - 110	P	10/23/2024	15:57	LB133086
	Lead	5020	5000	100	90 - 110	P	10/23/2024	15:57	LB133086
CCV02	Copper	1180	1250	94	90 - 110	P	10/23/2024	16:48	LB133086
	Iron	4660	5000	93	90 - 110	P	10/23/2024	16:48	LB133086
	Lead	4810	5000	96	90 - 110	P	10/23/2024	16:48	LB133086
CCV03	Copper	1190	1250	95	90 - 110	P	10/23/2024	17:42	LB133086
	Iron	4840	5000	97	90 - 110	P	10/23/2024	17:42	LB133086
	Lead	4850	5000	97	90 - 110	P	10/23/2024	17:42	LB133086
CCV04	Copper	1240	1250	99	90 - 110	P	10/23/2024	18:32	LB133086
	Iron	4820	5000	96	90 - 110	P	10/23/2024	18:32	LB133086
	Lead	5080	5000	102	90 - 110	P	10/23/2024	18:32	LB133086
CCV05	Copper	1200	1250	96	90 - 110	P	10/23/2024	19:23	LB133086
	Iron	4650	5000	93	90 - 110	P	10/23/2024	19:23	LB133086
	Lead	4910	5000	98	90 - 110	P	10/23/2024	19:23	LB133086
CCV06	Copper	1190	1250	95	90 - 110	P	10/23/2024	20:15	LB133086
	Iron	4690	5000	94	90 - 110	P	10/23/2024	20:15	LB133086
	Lead	4910	5000	98	90 - 110	P	10/23/2024	20:15	LB133086
CCV07	Copper	1170	1250	94	90 - 110	P	10/23/2024	21:28	LB133086
	Iron	5380	5000	108	90 - 110	P	10/23/2024	21:28	LB133086
	Lead	4820	5000	96	90 - 110	P	10/23/2024	21:28	LB133086
CCV08	Copper	1170	1250	94	90 - 110	P	10/23/2024	22:20	LB133086
	Iron	4550	5000	91	90 - 110	P	10/23/2024	22:20	LB133086
	Lead	4980	5000	100	90 - 110	P	10/23/2024	22:20	LB133086