



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	22.0	20.0	110	40 - 160	P	10/22/2024	15:24	LB133061
	Iron	101	100	102	40 - 160	P	10/22/2024	15:24	LB133061
	Lead	11.2	12.0	94	40 - 160	P	10/22/2024	15:24	LB133061
CRI01	Copper	20.9	20.0	105	40 - 160	P	10/23/2024	15:22	LB133086
	Iron	102	100	102	40 - 160	P	10/23/2024	15:22	LB133086
	Lead	11.1	12.0	92	40 - 160	P	10/23/2024	15:22	LB133086