



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tully Environmental, Inc				SDG No.:	P4464			
Contract:	TULL01	Lab Code:	CHEM	Case No.:	P4464	SAS No.:	P4464		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	15:20	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	15:20	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	15:20	LB133061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>P4464</u>							
Contract: <u>TULL01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4464</u>	SAS No.: <u>P4464</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	15:56	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	15:56	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	15:56	LB133061
CCB02	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	16:34	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	16:34	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	16:34	LB133061
CCB03	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	17:29	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	17:29	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	17:29	LB133061
CCB04	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	18:23	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	18:23	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	18:23	LB133061
CCB05	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	19:17	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	19:17	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	19:17	LB133061
CCB06	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	20:29	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	20:29	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	20:29	LB133061
CCB07	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	21:20	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	21:20	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	21:20	LB133061
CCB08	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	22:33	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	22:33	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	22:33	LB133061
CCB09	Copper	20.0	+/-20.0	U	20.0	P	10/22/2024	23:24	LB133061
	Iron	100	+/-100	U	100	P	10/22/2024	23:24	LB133061
	Lead	12.0	+/-12.0	U	12.0	P	10/22/2024	23:24	LB133061

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: P4464

Contract: TULL01

Lab Code: CHEM

Case No.: P4464

SAS No.: P4464

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	15:16	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	15:16	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	15:16	LB133086

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>P4464</u>							
Contract: <u>TULL01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4464</u>	SAS No.: <u>P4464</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	16:02	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	16:02	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	16:02	LB133086
CCB02	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	16:52	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	16:52	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	16:52	LB133086
CCB03	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	17:46	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	17:46	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	17:46	LB133086
CCB04	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	18:36	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	18:36	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	18:36	LB133086
CCB05	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	19:28	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	19:28	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	19:28	LB133086
CCB06	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	20:21	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	20:21	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	20:21	LB133086
CCB07	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	21:34	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	21:34	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	21:34	LB133086
CCB08	Copper	20.0	+/-20.0	U	20.0	P	10/23/2024	22:25	LB133086
	Iron	100	+/-100	U	100	P	10/23/2024	22:25	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	22:25	LB133086