

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** P4464
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P4464 **SAS No.:** P4464
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-0.20	2.0	10	-18	22	10/22/2024	15:29	LB133061
	Iron	104000	101000	103	85600	116500	10/22/2024	15:29	LB133061
	Lead	8.45			-12	12	10/22/2024	15:29	LB133061
ICSAB01	Copper	502	511	98	434	588	10/22/2024	15:33	LB133061
	Iron	103000	99300	104	84400	114500	10/22/2024	15:33	LB133061
	Lead	56.9	49.0	116	37	61	10/22/2024	15:33	LB133061
ICSA	Copper	3.05	2.0	152	-18	22	10/22/2024	15:38	LB133061
	Iron	110000	101000	109	85600	116500	10/22/2024	15:38	LB133061
	Lead	17.4			-12	12	10/22/2024	15:38	LB133061
ICSAB	Copper	535	511	105	434	588	10/22/2024	15:42	LB133061
	Iron	104000	99300	105	84400	114500	10/22/2024	15:42	LB133061
	Lead	73.1	49.0	149	37	61	10/22/2024	15:42	LB133061
ICSA01	Copper	2.08	2.0	104	-18	22	10/23/2024	15:26	LB133086
	Iron	102000	101000	101	85600	116500	10/23/2024	15:26	LB133086
	Lead	7.85			-12	12	10/23/2024	15:26	LB133086
ICSAB01	Copper	486	511	95	434	588	10/23/2024	15:32	LB133086
	Iron	100000	99300	101	84400	114500	10/23/2024	15:32	LB133086
	Lead	55.0	49.0	112	37	61	10/23/2024	15:32	LB133086
ICSA	Copper	12.5	2.0	625	-18	22	10/23/2024	15:36	LB133086
	Iron	99600	101000	99	85600	116500	10/23/2024	15:36	LB133086
	Lead	23.8			-12	12	10/23/2024	15:36	LB133086
ICSAB	Copper	522	511	102	434	588	10/23/2024	15:40	LB133086
	Iron	104000	99300	105	84400	114500	10/23/2024	15:40	LB133086
	Lead	74.7	49.0	152	37	61	10/23/2024	15:40	LB133086