

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>P4464</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>P4464</u> sas no.: <u>P4464</u>
matrix: <u>Water</u>	sample id: <u>P4347-01</u>	client id: <u>EFFLUENT-DAY-1-METALSMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4347-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	275		119		150	104		P
Iron	ug/L	75 - 125	1580		166		1500	94		P
Lead	ug/L	75 - 125	462		1.87	J	500	92		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>P4464</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>P4464</u> sas no.: <u>P4464</u>
matrix: <u>Water</u>	sample id: <u>P4347-01</u>	client id: <u>EFFLUENT-DAY-1-METALMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4347-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	263		119		150	96		P
Iron	ug/L	75 - 125	1560		166		1500	93		P
Lead	ug/L	75 - 125	447		1.87	J	500	89		P