

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: P4464

Contract: TULL01

Lab Code: CHEM

Case No.: P4464

SAS No.: P4464

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164302BS							
Copper	ug/L	150	145		97	85 - 115	P
Iron	ug/L	1500	1410		94	85 - 115	P
Lead	ug/L	500	465		93	85 - 115	P