

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100824\
 Data File : VN084343.D
 Acq On : 08 Oct 2024 12:09
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN100824

Quant Time: Oct 09 02:27:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:21:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	103	0.00
2 M	Dichlorodifluoromethane	1.832	1.811	1.1	97	0.00
3 M	Chloromethane	2.084	2.079	0.2	100	0.00
4 M	Vinyl Chloride	2.044	2.070	-1.3	102	0.00
5 M	Bromomethane	1.354	1.335	1.4	97	0.00
6 M	Chloroethane	1.340	1.370	-2.2	101	0.00
7 M	Trichlorofluoromethane	3.329	3.232	2.9	98	0.00
8 T	Diethyl Ether	1.224	1.196	2.3	100	0.00
9	1,1,2-Trichlorotrifluoroeth	1.872	1.922	-2.7	105	0.01
10 M	1,1-Dichloroethene	1.868	1.806	3.3	102	0.00
11	Methyl Iodide	2.455	2.363	3.7	99	0.00
12	Methyl Acetate	2.339	2.219	5.1	100	0.00
13 M	Acrolein	0.434	0.410	5.5	105	0.00
14 M	Acrylonitrile	1.051	0.975	7.2	94	0.00
15 M	Acetone	0.300	0.333	-11.0	120	0.00
16 M	Carbon Disulfide	5.562	5.286	5.0	96	0.00
17	Allyl chloride	3.028	2.886	4.7	96	0.00
18 M	Methylene Chloride	2.104	2.118	-0.7	102	0.00
19 M	trans-1,2-Dichloroethene	1.977	1.921	2.8	99	0.00
20 T	Diisopropyl ether	6.510	6.628	-1.8	103	0.00
21 M	1,1-Dichloroethane	3.736	3.726	0.3	100	0.00
22 M	cis-1,2-Dichloroethene	2.300	2.294	0.3	100	0.00
23 M	tert-Butyl Alcohol	0.403	0.361	10.4	94	0.00
24 M	Methyl tert-Butyl Ether	6.242	6.148	1.5	101	0.00
25 M	Chloroform	3.783	3.847	-1.7	102	0.00
26	Cyclohexane	3.124	2.957	5.3	93	0.00
27 s	1,2-Dichloroethane-d4	2.415	2.416	-0.0	101	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
29	1,1-Dichloropropene	0.492	0.493	-0.2	100	0.00
30 M	2-Butanone	0.262	0.264	-0.8	103	0.00
31	2,2-Dichloropropane	0.579	0.608	-5.0	106	0.00
32 M	1,1,1-Trichloroethane	0.629	0.647	-2.9	100	0.00
33 M	Carbon Tetrachloride	0.545	0.546	-0.2	101	0.00
34 M	Benzene	1.542	1.557	-1.0	100	0.00
35	Methacrylonitrile	0.280	0.273	2.5	100	0.00
36 M	1,2-Dichloroethane	0.514	0.522	-1.6	100	0.00
37 M	Trichloroethene	0.355	0.360	-1.4	99	0.00
38	Methylcyclohexane	0.541	0.536	0.9	103	0.00
39 M	1,2-Dichloropropane	0.376	0.384	-2.1	100	0.00
40	Dibromomethane	0.263	0.266	-1.1	101	0.00
41 M	Bromodichloromethane	0.558	0.560	-0.4	100	0.00
42 M	Vinyl Acetate	0.912	0.901	1.2	102	0.00
43	Ethyl Acetate	0.517	0.490	5.2	96	0.00
44	Isopropyl Acetate	0.846	0.822	2.8	98	0.00
45 T	1,4-Dioxane	0.008	0.007	12.5	96	0.00
46	Methyl methacrylate	0.409	0.396	3.2	98	0.00
47	n-amyl Acetate	0.753	0.731	2.9	99	0.00
48 M	t-1,3-Dichloropropene	0.576	0.579	-0.5	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100824\
 Data File : VN084343.D
 Acq On : 08 Oct 2024 12:09
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN100824

Quant Time: Oct 09 02:27:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100824W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 09 02:21:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.604	0.622	-3.0	102	0.00
50 M	1,1,2-Trichloroethane	0.353	0.354	-0.3	100	0.00
51	Ethyl methacrylate	0.583	0.573	1.7	100	0.00
52	1,3-Dichloropropane	0.617	0.620	-0.5	99	0.00
53 M	Dibromochloromethane	0.417	0.422	-1.2	102	0.00
54 M	1,2-Dibromoethane	0.364	0.355	2.5	96	0.00
55 M	2-Chloroethyl vinyl ether	0.265	0.233	12.1	93	0.00
56 M	Bromoform	0.277	0.278	-0.4	106	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
58 M	4-Methyl-2-Pentanone	0.559	0.544	2.7	97	0.00
59 M	2-Hexanone	0.415	0.413	0.5	101	0.00
60 S	4-Bromofluorobenzene	0.495	0.508	-2.6	103	0.00
61 M	Tetrachloroethene	0.339	0.347	-2.4	102	0.00
62 M	Toluene	1.756	1.805	-2.8	102	0.00
63 S	Toluene-d8	1.390	1.407	-1.2	101	0.00
64 M	Chlorobenzene	1.093	1.091	0.2	100	0.00
65	1,1,1,2-Tetrachloroethane	0.380	0.383	-0.8	100	0.00
66 M	Ethyl Benzene	1.908	1.912	-0.2	101	0.00
67 M	m/p-Xylenes	0.724	0.749	-3.5	103	0.00
68 M	o-Xylene	0.698	0.700	-0.3	101	0.00
69 M	Styrene	1.198	1.205	-0.6	102	0.00
70	Isopropylbenzene	1.781	1.774	0.4	99	0.00
71 M	1,1,2,2-Tetrachloroethane	0.551	0.549	0.4	100	0.00
72	1,2,3-Trichloropropane	0.472	0.532	-12.7	114	0.00
73	Bromobenzene	0.431	0.431	0.0	101	0.00
74	n-propylbenzene	2.123	2.172	-2.3	103	0.00
75	2-Chlorotoluene	1.319	1.319	0.0	100	0.00
76	1,3,5-Trimethylbenzene	1.500	1.514	-0.9	102	0.00
77	t-1,4-Dichloro-2-butene	0.202	0.190	5.9	102	0.00
78	4-Chlorotoluene	1.341	1.341	0.0	101	0.00
79	tert-butylbenzene	1.279	1.292	-1.0	102	0.00
80	1,2,4-Trimethylbenzene	1.522	1.552	-2.0	103	0.00
81	sec-Butylbenzene	1.762	1.768	-0.3	101	0.00
82	p-Isopropyltoluene	1.483	1.484	-0.1	103	0.00
83 M	1,3-Dichlorobenzene	0.802	0.800	0.2	101	0.00
84 M	1,4-Dichlorobenzene	0.796	0.772	3.0	101	0.00
85	n-Butylbenzene	1.284	1.234	3.9	103	0.00
86 T	Hexachloroethane	0.286	0.291	-1.7	104	0.00
87 M	1,2-Dichlorobenzene	0.775	0.760	1.9	100	0.00
88	1,2-Dibromo-3-Chloropropane	0.113	0.106	6.2	96	0.00
89	1,2,4-Trichlorobenzene	0.403	0.364	9.7	95	0.00
90	Hexachlorobutadiene	0.187	0.180	3.7	104	0.00
91 M	Naphthalene	1.366	1.147	16.0	89	0.00
92	1,2,3-Trichlorobenzene	0.403	0.364	9.7	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0