

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P4464 SAS No.: P4464 SDG No.: P4464

Instrument ID: MSVOA_N Calibration Date/Time: 10/23/2024 13:26

Lab File ID: VN084469.D Init. Calib. Date(s): 10/08/2024 10/08/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:43 11:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.542	1.545	0.5	0.19	
Toluene	1.756	1.718	0.4	-2.16	
Ethyl Benzene	1.908	1.782	0.1	-6.6	
m/p-Xylenes	0.724	0.700	0.3	-3.32	
o-Xylene	0.698	0.653	0.3	-6.45	
1,2-Dichloroethane-d4	2.415	2.435	0.01	0.83	
Toluene-d8	1.390	1.458	0.01	4.89	
4-Bromofluorobenzene	0.495	0.505	0.2	2.02	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.