

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN100824

Review By	John Carlone	Review On	10/9/2024 9:46:14 AM
Supervise By	Maresh Dadoda	Supervise On	10/9/2024 8:56:13 PM
SubDirectory	VN100824	HP Acquire Method	HP Processing Method 624N100824W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP130714		
Initial Calibration Stds	VP130724,VP130725,VP130726,VP130727,VP130728		
CCC	VP130715		
Internal Standard/PEM			
ICV/I.BLK	VP130729		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084336.D	08 Oct 2024 09:08		JC\MD	Ok
2	VSTDICC150	VSTDICC150	VN084337.D	08 Oct 2024 09:43	V13516	JC\MD	Ok,M
3	VSTDICC100	VSTDICC100	VN084338.D	08 Oct 2024 10:07		JC\MD	Ok,M
4	VSTDICC050	VSTDICC050	VN084339.D	08 Oct 2024 10:31		JC\MD	Ok,M
5	VSTDICCC020	VSTDICCC020	VN084340.D	08 Oct 2024 10:55		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN084341.D	08 Oct 2024 11:19		JC\MD	Ok,M
7	IBLK	IBLK	VN084342.D	08 Oct 2024 11:43		JC\MD	Ok
8	VSTDICV020	ICVVN100824	VN084343.D	08 Oct 2024 12:09		JC\MD	Ok,M
9	VN1008WBS01	VN1008WBS01	VN084344.D	08 Oct 2024 12:46		JC\MD	Ok,M
10	VN1008WBSD01	VN1008WBSD01	VN084345.D	08 Oct 2024 13:21		JC\MD	Ok,M
11	VN1008WBL01	VN1008WBL01	VN084346.D	08 Oct 2024 13:45		JC\MD	Ok
12	P4271-01	14B-1	VN084347.D	08 Oct 2024 14:11	vial B pH#6.0	JC\MD	Ok
13	P4334-01	001-WILLETS-PT-BLVD	VN084348.D	08 Oct 2024 14:53	vial A pH<2	JC\MD	Ok,M
14	P4334-02	002-35TH-AVE(SEP)	VN084349.D	08 Oct 2024 15:17	vial A pH<2	JC\MD	Ok
15	P4337-01	001-WILLETS-PT-BLVD	VN084350.D	08 Oct 2024 15:41	vial A pH<2	JC\MD	Ok
16	P4337-02	002-35TH-AVE(OCT)	VN084351.D	08 Oct 2024 16:05	vial A pH<2	JC\MD	Ok
17	P4289-01	EFF-Waste Water	VN084352.D	08 Oct 2024 16:29	vial B pH<2	JC\MD	Ok,M
18	IBLK	IBLK	VN084353.D	08 Oct 2024 16:53		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN100824

Review By	John Carlone	Review On	10/9/2024 9:46:14 AM
Supervise By	Mahesh Dadoda	Supervise On	10/9/2024 8:56:13 PM
SubDirectory	VN100824	HP Acquire Method	HP Processing Method 624N100824W.M

STD. NAME	STD REF.#
Tune/Reschk	VP130714
Initial Calibration Stds	VP130724,VP130725,VP130726,VP130727,VP130728
CCC	VP130715
Internal Standard/PEM	
ICV/I.BLK	VP130729
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	IBLK	IBLK	VN084354.D	08 Oct 2024 17:16		JC\MD	Ok,M
20	VSTDCCC020	VSTDCCC020EC	VN084355.D	08 Oct 2024 17:40		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102324

Review By	Semsettin Yesilyurt	Review On	10/24/2024 9:12:55 AM
Supervise By	Mahesh Dadoda	Supervise On	10/24/2024 2:47:12 PM
SubDirectory	VN102324	HP Acquire Method	MSVOA_N HP Processing Method 624N100824W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131051
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131052,VP131053

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084465.D	23 Oct 2024 09:13	Not used	JC\MD	Not Ok
2	VSTDCCC050	VSTDCCC050	VN084466.D	23 Oct 2024 10:01	Not used	JC\MD	Not Ok
3	VN1023MBL01	VN1023MBL01	VN084467.D	23 Oct 2024 11:24	Not used	JC\MD	Not Ok
4	BFB	BFB	VN084468.D	23 Oct 2024 12:39		JC\MD	Ok
5	VSTDCCC020	VSTDCCC020	VN084469.D	23 Oct 2024 13:26		JC\MD	Ok,M
6	VN1023WBS01	VN1023WBS01	VN084470.D	23 Oct 2024 13:50		JC\MD	Ok,M
7	VN1023WBSD01	VN1023WBSD01	VN084471.D	23 Oct 2024 14:14		JC\MD	Ok,M
8	VN1023WBL01	VN1023WBL01	VN084472.D	23 Oct 2024 15:02		JC\MD	Ok
9	P4464-01	001-WILLETS-PT-BLVD	VN084473.D	23 Oct 2024 15:26	pH#1.0 A	JC\MD	Ok
10	VSTDCCC020	VSTDCCC020EC	VN084474.D	23 Oct 2024 16:10	Bad purge	JC\MD	Not Ok

M : Manual Integration