

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4464

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VN084471.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1023WBSD01	Benzene	20	19.8	ug/L	99	4		65	135	20
	Toluene	20	19.5	ug/L	98	1		70	130	20
	Ethyl Benzene	20	18.8	ug/L	94	3		60	140	20
	m/p-Xylenes	40	37.8	ug/L	95	1		87	111	20
	o-Xylene	20	18.8	ug/L	94	3		87	111	20