

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: P4464 SAS No.: P4464 SDG NO.: P4464
 Lab File ID: VN084469.D Date Analyzed: 10/23/2024
 Instrument ID: MSVOA_N Time Analyzed: 13:26
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	34138	7.81	175440	9.10	166908	11.87
UPPER LIMIT	68276	8.312	350880	9.6	333816	12.365
LOWER LIMIT	17069	7.312	87720	8.6	83454	11.365
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (SEP)	30612	7.81	143930	9.10	140820	11.87
VN1023WBL01	29639	7.82	140547	9.10	131687	11.87
VN1023WBS01	34001	7.81	177325	9.10	165233	11.87
VN1023WBSD01	31512	7.82	161510	9.10	152709	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.