

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: P4464 SAS No.: P4464 SDG NO.: P4464
Lab File ID: VN084336.D BFB Injection Date: 10/08/2024
Instrument ID: MSVOA_N BFB Injection Time: 09:08
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19
75	30.0 - 60.0% of mass 95	51.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	73.7
175	5.0 - 9.0% of mass 174	5.5 (7.5) 1
176	95.0 - 101.0% of mass 174	70.3 (95.4) 1
177	5.0 - 9.0% of mass 176	4.4 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC150	VSTDICC150	VN084337.D	10/08/2024	09:43
VSTDICC100	VSTDICC100	VN084338.D	10/08/2024	10:07
VSTDICC050	VSTDICC050	VN084339.D	10/08/2024	10:31
VSTDICCC020	VSTDICCC020	VN084340.D	10/08/2024	10:55
VSTDICC005	VSTDICC005	VN084341.D	10/08/2024	11:19