



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Portal Partners Tri-Venture**

**Project Name: Amtrak Sawtooth Bridges 2024**

**Project # N/A**

**Chemtech Project # P4471**

**Test Name: Hexavalent Chromium, Trivalent Chromium**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 10/21/2024.

1 Water sample was received on 10/21/2024.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: EPH, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, PCB, SVOC-TCL BNA -20, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for Hexavalent Chromium, Trivalent Chromium.

### **C. Analytical Techniques:**

The analysis of Trivalent Chromium was based on method 6010D and The analysis of Hexavalent Chromium was based on method 7196A.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

The temperature of the samples at the time of receipt was 23.0°C. Lab notified this issue to the client. See the communication in shipping Document section.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_