



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** Portal Partners Tri-Venture  
**Project:** Amtrak Sawtooth Bridges 2024

**SDG No.:** P4471  
**RunNo.:** LB133044

| Analyte                                       | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-----------------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>Hexavalent Chromium  | mg/L  | 0.503  | 0.5        | 101        | 90-110                 | 10/22/2024    |
| Sample ID: <b>CCV1</b><br>Hexavalent Chromium | mg/L  | 0.499  | 0.5        | 100        | 90-110                 | 10/22/2024    |
| Sample ID: <b>CCV2</b><br>Hexavalent Chromium | mg/L  | 0.500  | 0.5        | 100        | 90-110                 | 10/22/2024    |
| Sample ID: <b>CCV3</b><br>Hexavalent Chromium | mg/L  | 0.503  | 0.5        | 101        | 90-110                 | 10/22/2024    |