

Continuing Calibration Report for SequenceID : FD102224AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048560.D

Aromatic C10-C12	8157685.000	40.000	4.096	5.814	203942.125	180984.840	-12.685
Aromatic C12-C16	12219474.000	60.000	5.815	8.420	203657.900	189995.313	-7.191
Aromatic C16-C21	14694945.000	80.000	8.421	12.681	183686.813	178728.230	-2.774
Aromatic C21-C36	31489467.000	180.000	12.682	18.086	174941.483	165360.750	-5.794
Aromatic EPH	66561571.000	360.000	4.096	18.086	184893.253	174173.072	-6.155