

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG NO.: P4471

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068151.D Time Analyzed: 09:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.923	5.819	6.019	509.490	500.000	1.9
Aroclor-1016-2	5.945	5.842	6.042	522.480	500.000	4.5
Aroclor-1016-3	6.009	5.905	6.105	515.710	500.000	3.1
Aroclor-1016-4	6.107	6.003	6.203	506.700	500.000	1.3
Aroclor-1016-5	6.401	6.298	6.498	491.580	500.000	-1.7
Aroclor-1260-1	7.525	7.422	7.622	469.920	500.000	-6.0
Aroclor-1260-2	7.778	7.675	7.875	469.080	500.000	-6.2
Aroclor-1260-3	8.139	8.037	8.237	478.230	500.000	-4.4
Aroclor-1260-4	8.377	8.275	8.475	470.780	500.000	-5.8
Aroclor-1260-5	8.714	8.612	8.812	474.520	500.000	-5.1
Decachlorobiphenyl	10.672	10.569	10.769	49.780	50.000	-0.4
Tetrachloro-m-xylene	4.759	4.655	4.855	52.360	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG NO.: P4471

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068151.D Time Analyzed: 09:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.159	5.061	5.261	471.440	500.000	-5.7
Aroclor-1016-2	5.178	5.081	5.281	472.270	500.000	-5.5
Aroclor-1016-3	5.359	5.262	5.462	475.440	500.000	-4.9
Aroclor-1016-4	5.398	5.301	5.501	471.380	500.000	-5.7
Aroclor-1016-5	5.617	5.520	5.720	472.660	500.000	-5.5
Aroclor-1260-1	6.660	6.564	6.764	464.670	500.000	-7.1
Aroclor-1260-2	6.846	6.750	6.950	462.090	500.000	-7.6
Aroclor-1260-3	7.004	6.908	7.108	468.280	500.000	-6.3
Aroclor-1260-4	7.479	7.383	7.583	468.800	500.000	-6.2
Aroclor-1260-5	7.718	7.622	7.822	466.860	500.000	-6.6
Decachlorobiphenyl	9.211	9.119	9.319	47.830	50.000	-4.3
Tetrachloro-m-xylene	4.052	3.952	4.152	48.140	50.000	-3.7