

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG NO.: P4471

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068166.D Time Analyzed: 15:57

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.921  | 5.819     | 6.019  | 491.140            | 500.000           | -1.8  |
| Aroclor-1016-2       | 5.944  | 5.842     | 6.042  | 505.430            | 500.000           | 1.1   |
| Aroclor-1016-3       | 6.007  | 5.905     | 6.105  | 491.780            | 500.000           | -1.6  |
| Aroclor-1016-4       | 6.105  | 6.003     | 6.203  | 485.430            | 500.000           | -2.9  |
| Aroclor-1016-5       | 6.399  | 6.298     | 6.498  | 476.670            | 500.000           | -4.7  |
| Aroclor-1260-1       | 7.523  | 7.422     | 7.622  | 449.530            | 500.000           | -10.1 |
| Aroclor-1260-2       | 7.776  | 7.675     | 7.875  | 448.260            | 500.000           | -10.3 |
| Aroclor-1260-3       | 8.138  | 8.037     | 8.237  | 457.380            | 500.000           | -8.5  |
| Aroclor-1260-4       | 8.376  | 8.275     | 8.475  | 456.620            | 500.000           | -8.7  |
| Aroclor-1260-5       | 8.713  | 8.612     | 8.812  | 462.630            | 500.000           | -7.5  |
| Decachlorobiphenyl   | 10.670 | 10.569    | 10.769 | 48.800             | 50.000            | -2.4  |
| Tetrachloro-m-xylene | 4.757  | 4.655     | 4.855  | 51.470             | 50.000            | 2.9   |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG NO.: P4471

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068166.D Time Analyzed: 15:57

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 5.157 | 5.061     | 5.261 | 465.090            | 500.000           | -7.0 |
| Aroclor-1016-2       | 5.177 | 5.081     | 5.281 | 473.530            | 500.000           | -5.3 |
| Aroclor-1016-3       | 5.357 | 5.262     | 5.462 | 476.580            | 500.000           | -4.7 |
| Aroclor-1016-4       | 5.397 | 5.301     | 5.501 | 471.310            | 500.000           | -5.7 |
| Aroclor-1016-5       | 5.616 | 5.520     | 5.720 | 467.820            | 500.000           | -6.4 |
| Aroclor-1260-1       | 6.659 | 6.564     | 6.764 | 451.860            | 500.000           | -9.6 |
| Aroclor-1260-2       | 6.846 | 6.750     | 6.950 | 457.130            | 500.000           | -8.6 |
| Aroclor-1260-3       | 7.004 | 6.908     | 7.108 | 455.780            | 500.000           | -8.8 |
| Aroclor-1260-4       | 7.478 | 7.383     | 7.583 | 458.570            | 500.000           | -8.3 |
| Aroclor-1260-5       | 7.716 | 7.622     | 7.822 | 455.920            | 500.000           | -8.8 |
| Decachlorobiphenyl   | 9.209 | 9.119     | 9.319 | 46.300             | 50.000            | -7.4 |
| Tetrachloro-m-xylene | 4.050 | 3.952     | 4.152 | 48.480             | 50.000            | -3.0 |