

Surrogate Summary

SDG No.: **P4471**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP067586.D	PIBLK-PP067586.D	Tetrachloro-m-xylene	1	20	21.1	105		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.5	117		70 (60)	130 (140)
I.BLK-PP068124.D	PIBLK-PP068124.D	Tetrachloro-m-xylene	1	20	21.1	106		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.1	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	18.8	94		70 (60)	130 (140)
PB164310BL	PB164310BL	Tetrachloro-m-xylene	1	20	20.1	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.2	86		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.0	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.6	93		30 (32)	150 (175)
PB164310BS	PB164310BS	Tetrachloro-m-xylene	1	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.5	88		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.7	94		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.5	93		30 (32)	150 (175)
P4471-01	B-180-SB01	Tetrachloro-m-xylene	1	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.3	82		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.3	91		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.6	73		30 (32)	150 (175)
I.BLK-PP068139.D	PIBLK-PP068139.D	Tetrachloro-m-xylene	1	20	20.0	100		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.4	102		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.2	96		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	19.5	98		70 (60)	130 (140)
P4472-05MS	BP-F-6MS	Tetrachloro-m-xylene	1	20	21.0	105		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.1	100		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.3	91		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.6	93		30 (32)	150 (175)
P4472-05MSD	BP-F-6MSD	Tetrachloro-m-xylene	1	20	21.3	106		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.1	101		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.2	91		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.8	94		30 (32)	150 (175)
I.BLK-PP068149.D	PIBLK-PP068149.D	Tetrachloro-m-xylene	1	20	20.8	104		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	22.1	111		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.0	95		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.2	101		70 (60)	130 (140)
I.BLK-PP068155.D	PIBLK-PP068155.D	Tetrachloro-m-xylene	1	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	22.2	111		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.7	99		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.9	105		70 (60)	130 (140)
P4471-02	B-180-SB02	Tetrachloro-m-xylene	1	20	11.3	56		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: P4471

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4471-02	B-180-SB02	Decachlorobiphenyl	1	20	7.13	36		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	10.9	54		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	7.51	38		30 (32)	150 (175)
I.BLK-PP068170.D	PIBLK-PP068170.D	Tetrachloro-m-xylene	1	20	20.8	104		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.9	110		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.5	98		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.9	104		70 (60)	130 (140)