

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06  
Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG No.: P4471  
Instrument ID: MSVOA\_N Calibration Date(s): 09/30/2024 09/30/2024  
Heated Purge: (Y/N) N Calibration Time(s): 12:25 14:48  
GC Column: RXI-624 ID: 0.25 (mm)

|                                |                     |        |                     |        |                     |        |                     |       |                     |  |                     |  |
|--------------------------------|---------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|---------------------|--|---------------------|--|
| LAB FILE ID:                   | RRF100 = VN084213.D |        | RRF050 = VN084214.D |        | RRF020 = VN084215.D |        | RRF010 = VN084216.D |       | RRF005 = VN084217.D |  | RRF001 = VN084218.D |  |
| COMPOUND                       | RRF100              | RRF050 | RRF020              | RRF010 | RRF005              | RRF001 | RRF                 | % RSD |                     |  |                     |  |
| Dichlorodifluoromethane        | 0.567               | 0.613  | 0.600               | 0.562  | 0.607               | 0.600  | 0.591               | 3.7   |                     |  |                     |  |
| Chloromethane                  | 0.619               | 0.671  | 0.677               | 0.648  | 0.747               | 0.690  | 0.675               | 6.4   |                     |  |                     |  |
| Vinyl Chloride                 | 0.611               | 0.667  | 0.665               | 0.641  | 0.724               | 0.617  | 0.654               | 6.4   |                     |  |                     |  |
| Bromomethane                   | 0.368               | 0.432  | 0.432               | 0.424  | 0.501               |        | 0.431               | 10.9  |                     |  |                     |  |
| Chloroethane                   | 0.375               | 0.419  | 0.430               | 0.433  | 0.522               | 0.632  | 0.468               | 19.9  |                     |  |                     |  |
| Trichlorofluoromethane         | 0.953               | 1.060  | 1.047               | 0.959  | 1.113               | 0.978  | 1.019               | 6.4   |                     |  |                     |  |
| 1,1,2-Trichlorotrifluoroethane | 0.542               | 0.590  | 0.603               | 0.532  | 0.633               | 0.602  | 0.584               | 6.7   |                     |  |                     |  |
| 1,1-Dichloroethene             | 0.538               | 0.587  | 0.596               | 0.533  | 0.631               | 0.493  | 0.563               | 8.9   |                     |  |                     |  |
| Acetone                        | 0.299               | 0.342  | 0.337               | 0.298  | 0.336               | 0.299  | 0.318               | 6.8   |                     |  |                     |  |
| Carbon Disulfide               | 1.588               | 1.723  | 1.746               | 1.650  | 1.908               | 2.080  | 1.782               | 10.2  |                     |  |                     |  |
| Methyl tert-butyl Ether        | 1.825               | 2.033  | 2.035               | 1.802  | 2.049               | 1.656  | 1.900               | 8.6   |                     |  |                     |  |
| Methyl Acetate                 | 0.607               | 0.707  | 0.694               | 0.691  | 0.765               | 0.700  | 0.694               | 7.3   |                     |  |                     |  |
| Methylene Chloride             | 0.594               | 0.655  | 0.661               | 0.618  | 0.669               | 0.686  | 0.647               | 5.3   |                     |  |                     |  |
| trans-1,2-Dichloroethene       | 0.555               | 0.622  | 0.619               | 0.570  | 0.627               | 0.546  | 0.590               | 6.2   |                     |  |                     |  |
| 1,1-Dichloroethane             | 1.075               | 1.193  | 1.163               | 1.084  | 1.226               | 1.046  | 1.131               | 6.4   |                     |  |                     |  |
| Cyclohexane                    | 0.970               | 1.068  | 1.084               | 1.047  | 1.321               |        | 1.098               | 12    |                     |  |                     |  |
| 2-Butanone                     | 0.395               | 0.452  | 0.465               | 0.419  | 0.467               | 0.404  | 0.434               | 7.3   |                     |  |                     |  |
| Carbon Tetrachloride           | 0.515               | 0.549  | 0.553               | 0.508  | 0.545               | 0.482  | 0.525               | 5.4   |                     |  |                     |  |
| cis-1,2-Dichloroethene         | 0.670               | 0.741  | 0.741               | 0.655  | 0.767               | 0.703  | 0.713               | 6.2   |                     |  |                     |  |
| Bromochloromethane             | 0.472               | 0.494  | 0.486               | 0.423  | 0.619               | 0.535  | 0.505               | 13.2  |                     |  |                     |  |
| Chloroform                     | 1.083               | 1.204  | 1.205               | 1.117  | 1.259               | 1.181  | 1.175               | 5.5   |                     |  |                     |  |
| 1,1,1-Trichloroethane          | 0.997               | 1.102  | 1.089               | 1.018  | 1.154               | 0.972  | 1.055               | 6.7   |                     |  |                     |  |
| Methylcyclohexane              | 0.567               | 0.583  | 0.577               | 0.507  | 0.534               | 0.428  | 0.533               | 11    |                     |  |                     |  |
| Benzene                        | 1.434               | 1.553  | 1.559               | 1.410  | 1.574               | 1.421  | 1.492               | 5.2   |                     |  |                     |  |
| 1,2-Dichloroethane             | 0.480               | 0.528  | 0.524               | 0.498  | 0.544               | 0.460  | 0.506               | 6.3   |                     |  |                     |  |
| Trichloroethene                | 0.334               | 0.362  | 0.361               | 0.329  | 0.379               | 0.325  | 0.348               | 6.3   |                     |  |                     |  |
| 1,2-Dichloropropane            | 0.346               | 0.375  | 0.380               | 0.339  | 0.388               | 0.289  | 0.353               | 10.4  |                     |  |                     |  |
| Bromodichloromethane           | 0.521               | 0.557  | 0.556               | 0.497  | 0.570               | 0.475  | 0.529               | 7.2   |                     |  |                     |  |
| 4-Methyl-2-Pentanone           | 0.449               | 0.508  | 0.510               | 0.468  | 0.484               | 0.387  | 0.468               | 9.9   |                     |  |                     |  |
| Toluene                        | 0.904               | 0.970  | 0.963               | 0.861  | 0.920               | 0.840  | 0.910               | 5.8   |                     |  |                     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG No.: P4471

Instrument ID: MSVOA\_N Calibration Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Calibration Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        | RRF100 = VN084213.D |        | RRF050 = VN084214.D |        | RRF020 = VN084215.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF010 = VN084216.D |        | RRF005 = VN084217.D |        | RRF001 = VN084218.D |       |       |  |
| COMPOUND                    | RRF100 | RRF050              | RRF020 | RRF010              | RRF005 | RRF001              | RRF   | % RSD |  |
| t-1,3-Dichloropropene       | 0.562  | 0.590               | 0.573  | 0.517               | 0.564  | 0.430               | 0.539 | 10.9  |  |
| cis-1,3-Dichloropropene     | 0.592  | 0.638               | 0.608  | 0.569               | 0.606  | 0.469               | 0.580 | 10.2  |  |
| 1,1,2-Trichloroethane       | 0.317  | 0.349               | 0.347  | 0.316               | 0.340  | 0.288               | 0.326 | 7.3   |  |
| 2-Hexanone                  | 0.341  | 0.387               | 0.387  | 0.340               | 0.357  | 0.285               | 0.349 | 10.8  |  |
| Dibromochloromethane        | 0.384  | 0.418               | 0.409  | 0.374               | 0.396  | 0.330               | 0.385 | 8.1   |  |
| 1,2-Dibromoethane           | 0.328  | 0.358               | 0.362  | 0.315               | 0.361  | 0.309               | 0.339 | 7.2   |  |
| Tetrachloroethene           | 0.323  | 0.359               | 0.351  | 0.338               | 0.373  | 0.346               | 0.348 | 5     |  |
| Chlorobenzene               | 1.068  | 1.170               | 1.143  | 1.069               | 1.173  | 1.028               | 1.109 | 5.5   |  |
| Ethyl Benzene               | 1.988  | 2.121               | 2.028  | 1.840               | 2.030  | 1.756               | 1.961 | 6.9   |  |
| m/p-Xylenes                 | 0.741  | 0.806               | 0.779  | 0.695               | 0.729  | 0.600               | 0.725 | 10    |  |
| o-Xylene                    | 0.714  | 0.774               | 0.738  | 0.666               | 0.734  | 0.491               | 0.686 | 14.9  |  |
| Styrene                     | 1.234  | 1.312               | 1.238  | 1.112               | 1.159  | 0.918               | 1.162 | 11.9  |  |
| Bromoform                   | 0.282  | 0.315               | 0.303  | 0.260               | 0.288  | 0.239               | 0.281 | 10    |  |
| Isopropylbenzene            | 3.737  | 4.132               | 4.055  | 3.677               | 3.864  | 3.428               | 3.815 | 6.8   |  |
| 1,1,2,2-Tetrachloroethane   | 1.001  | 1.183               | 1.187  | 1.127               | 1.291  | 1.095               | 1.147 | 8.5   |  |
| 1,3-Dichlorobenzene         | 1.624  | 1.787               | 1.780  | 1.646               | 1.843  | 1.679               | 1.727 | 5.1   |  |
| 1,4-Dichlorobenzene         | 1.628  | 1.784               | 1.734  | 1.638               | 1.889  | 1.789               | 1.744 | 5.7   |  |
| 1,2-Dichlorobenzene         | 1.555  | 1.710               | 1.740  | 1.613               | 1.769  | 1.770               | 1.693 | 5.3   |  |
| 1,2-Dibromo-3-Chloropropane | 0.209  | 0.238               | 0.253  | 0.242               | 0.271  | 0.207               | 0.237 | 10.4  |  |
| 1,2,4-Trichlorobenzene      | 0.859  | 0.933               | 0.923  | 0.792               | 0.839  | 0.657               | 0.834 | 12.2  |  |
| 1,2,3-Trichlorobenzene      | 0.816  | 0.896               | 0.903  | 0.820               | 0.822  | 0.751               | 0.835 | 6.8   |  |
| 1,2-Dichloroethane-d4       | 0.673  | 0.737               | 0.764  | 0.712               | 0.821  |                     | 0.741 | 7.5   |  |
| Dibromofluoromethane        | 0.308  | 0.324               | 0.341  | 0.316               | 0.359  |                     | 0.330 | 6.1   |  |
| Toluene-d8                  | 1.189  | 1.243               | 1.242  | 1.148               | 1.242  |                     | 1.213 | 3.5   |  |
| 4-Bromofluorobenzene        | 0.447  | 0.452               | 0.449  | 0.406               | 0.456  |                     | 0.442 | 4.6   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.