

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN093024

Review By	John Carlone	Review On	10/1/2024 9:40:01 AM
Supervise By	Maresh Dadoda	Supervise On	10/1/2024 5:34:39 PM
SubDirectory	VN093024	HP Acquire Method	HP Processing Method 82N093024W.M

STD. NAME	STD REF.#
Tune/Reschk	VP130570
Initial Calibration Stds	VP130580,VP130582,VP130583,VP130584,VP130585,VP130586
CCC	VP130571,VP130572
Internal Standard/PEM	
ICV/I.BLK	VP130587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084211.D	30 Sep 2024 09:24		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN084212.D	30 Sep 2024 11:45	Need ICAL	JC\MD	Not Ok
3	VSTDICC100	VSTDICC100	VN084213.D	30 Sep 2024 12:25		JC\MD	Ok,M
4	VSTDICCC050	VSTDICCC050	VN084214.D	30 Sep 2024 12:49		JC\MD	Ok,M
5	VSTDICC020	VSTDICC020	VN084215.D	30 Sep 2024 13:13		JC\MD	Ok,M
6	VSTDICC010	VSTDICC010	VN084216.D	30 Sep 2024 13:37		JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN084217.D	30 Sep 2024 14:00		JC\MD	Ok,M
8	VSTDICC001	VSTDICC001	VN084218.D	30 Sep 2024 14:48		JC\MD	Ok,M
9	IBLK	IBLK	VN084219.D	30 Sep 2024 15:12		JC\MD	Ok
10	VSTDICV050	ICVVN093024	VN084220.D	30 Sep 2024 15:36		JC\MD	Ok,M
11	VN0930MBL01	VN0930MBL01	VN084221.D	30 Sep 2024 16:11		JC\MD	Ok
12	VN0930WBL01	VN0930WBL01	VN084222.D	30 Sep 2024 16:35		JC\MD	Ok
13	VN0930WBS01	VN0930WBS01	VN084223.D	30 Sep 2024 17:42		JC\MD	Ok,M
14	VN0930WBSD01	VN0930WBSD01	VN084224.D	30 Sep 2024 18:06		JC\MD	Ok,M
15	P4116-24	RE132D5-20240918	VN084225.D	30 Sep 2024 18:30	vial A pH<2 Need 2x	JC\MD	Dilution
16	P4116-25	RE132D6-20240918	VN084226.D	30 Sep 2024 18:54	vial A pH<2 Need 40X	JC\MD	Dilution
17	P4116-26	RE132D7-20240918	VN084227.D	30 Sep 2024 19:18	vial A pH<2 Need 40X	JC\MD	Dilution
18	P4116-27	DUP05-20240918	VN084228.D	30 Sep 2024 19:42	vial A pH<2 Need 2x	JC\MD	Dilution

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN093024

Review By	John Carlone	Review On	10/1/2024 9:40:01 AM
Supervise By	Mahesh Dadoda	Supervise On	10/1/2024 5:34:39 PM
SubDirectory	VN093024	HP Acquire Method	HP Processing Method 82N093024W.M

STD. NAME	STD REF.#
Tune/Reschk	VP130570
Initial Calibration Stds	VP130580,VP130582,VP130583,VP130584,VP130585,VP130586
CCC	VP130571,VP130572
Internal Standard/PEM	
ICV/I.BLK	VP130587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDCCC050	VSTDCCC050EC	VN084229.D	30 Sep 2024 20:06		JC\MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102124

Review By	John Carlone	Review On	10/22/2024 9:50:25 AM
Supervise By	Maresh Dadoda	Supervise On	10/22/2024 3:37:17 PM
SubDirectory	VN102124	HP Acquire Method	HP Processing Method 82N093024W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP130983
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130986,VP130987

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084427.D	21 Oct 2024 11:14		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN084428.D	21 Oct 2024 11:38	V13516	JC\MD	Ok,M
3	VN1021MBL01	VN1021MBL01	VN084429.D	21 Oct 2024 12:12		JC\MD	Ok
4	VN1021WBL01	VN1021WBL01	VN084430.D	21 Oct 2024 12:36		JC\MD	Ok
5	VN1021MBS01	VN1021MBS01	VN084431.D	21 Oct 2024 13:00	Recovery failed	JC\MD	Not Ok
6	P4417-04	HOPPER-B-SOLID	VN084432.D	21 Oct 2024 13:24		JC\MD	Ok
7	VN1021WBS01	VN1021WBS01	VN084433.D	21 Oct 2024 13:48		JC\MD	Ok,M
8	VN1021WBSD01	VN1021WBSD01	VN084434.D	21 Oct 2024 14:22		JC\MD	Ok,M
9	P4438-03	VPB190-HYD-2024101	VN084435.D	21 Oct 2024 14:53	vial B pH<2	JC\MD	Ok
10	P4438-01	BP-VPB-190-TB-20241	VN084436.D	21 Oct 2024 15:17	vial B pH<2 TB	JC\MD	Ok
11	P4440-02	BPOW6-11-TB-202410	VN084437.D	21 Oct 2024 15:41	vial B pH<2 TB	JC\MD	Ok
12	P4438-02	BP-VPB-190-EB-20241	VN084438.D	21 Oct 2024 16:05	vial B pH<2 EB	JC\MD	Ok
13	P4438-04	BP-VPB-190-GW-58-60	VN084439.D	21 Oct 2024 16:29	vial B pH<2	JC\MD	Ok
14	P4438-05	BP-VPB-190-GW-98-10	VN084440.D	21 Oct 2024 16:53	vial B pH<2	JC\MD	Ok
15	P4438-06	BP-VPB-190-GW-158-1	VN084441.D	21 Oct 2024 17:17	vial B pH<2	JC\MD	Ok
16	P4438-07	BP-VPB-190-GW-198-2	VN084442.D	21 Oct 2024 17:40	vial B pH<2	JC\MD	Ok
17	P4440-01	BPOW6-11-HYD-20241	VN084443.D	21 Oct 2024 18:04	vial B pH<2	JC\MD	Ok
18	P4471-03	TB100824	VN084444.D	21 Oct 2024 18:28	vial A pH<2 TB	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102124

Review By	John Carlone	Review On	10/22/2024 9:50:25 AM
Supervise By	Maresh Dadoda	Supervise On	10/22/2024 3:37:17 PM
SubDirectory	VN102124	HP Acquire Method	HP Processing Method 82N093024W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP130983		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130986,VP130987		

19	P4460-05	TB-10182024	VN084445.D	21 Oct 2024 18:52	vial A pH<2 TB	JC\MD	Ok
20	P4460-06	WB-303-SW	VN084446.D	21 Oct 2024 19:16	vial A pH<2	JC\MD	Ok
21	VN1021MBS02	VN1021MBS02	VN084447.D	21 Oct 2024 19:40		JC\MD	Ok,M
22	VSTDCCC050	VSTDCCC050EC	VN084448.D	21 Oct 2024 20:04		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100924

Review By	Mahesh Dadoda	Review On	10/10/2024 10:09:46 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/10/2024 10:11:49 PM
SubDirectory	VY100924	HP Acquire Method	MSVOA_Y HP Processing Method 82y100924s.m

STD. NAME	STD REF.#
Tune/Reschk	VP130736
Initial Calibration Stds	VP130737,VP130738,VP130739,VP130740,VP130741,VP130742
CCC	VP130744,VP130745
Internal Standard/PEM	VP128297
ICV/I.BLK	VP130746
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY019826.D	09 Oct 2024 09:33		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY019827.D	09 Oct 2024 10:18		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY019828.D	09 Oct 2024 10:41		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY019829.D	09 Oct 2024 11:04		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY019830.D	09 Oct 2024 11:26	Comp. #11 is on Linear Regression	SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY019831.D	09 Oct 2024 11:49		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY019832.D	09 Oct 2024 12:11		SY/MD	Ok,M
8	VIBLK	VIBLK	VY019833.D	09 Oct 2024 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY100924	VY019834.D	09 Oct 2024 15:25		SY/MD	Ok,M
10	VY1009SBL01	VY1009SBL01	VY019835.D	09 Oct 2024 16:11		SY/MD	Ok
11	VY1009SBS01	VY1009SBS01	VY019836.D	09 Oct 2024 16:41		SY/MD	Ok,M
12	VY1009SBSD01	VY1009SBSD01	VY019837.D	09 Oct 2024 17:04		SY/MD	Ok,M
13	P4343-01	EO-03-100724	VY019838.D	09 Oct 2024 17:27	vial-B Not purged	SY/MD	Not Ok
14	P4353-02	CF-400-VOC-43	VY019839.D	09 Oct 2024 17:51	vial-A	SY/MD	Ok
15	P4359-01	EXCAVATION-SOIL	VY019840.D	09 Oct 2024 18:14	vial-A	SY/MD	Ok
16	P4344-01	TR-05-100724	VY019841.D	09 Oct 2024 18:38	vial-B	SY/MD	Ok
17	VSTDCC050	VSTDCC050EC	VY019842.D	09 Oct 2024 19:00		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102224

Review By	Maresh Dadoda	Review On	10/23/2024 1:02:15 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/24/2024 6:36:46 AM
SubDirectory	VY102224	HP Acquire Method	HP Processing Method 82y100924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131009
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131010,VP131011 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY019968.D	22 Oct 2024 08:49		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY019969.D	22 Oct 2024 10:01		SY/MD	Ok,M
3	VY1022SBL01	VY1022SBL01	VY019970.D	22 Oct 2024 10:36		SY/MD	Ok
4	VY1022SBS01	VY1022SBS01	VY019971.D	22 Oct 2024 11:16		SY/MD	Ok,M
5	VY1022SBSD01	VY1022SBSD01	VY019972.D	22 Oct 2024 11:38		SY/MD	Ok,M
6	P4460-03	WB-303-BOT	VY019973.D	22 Oct 2024 12:02	vial-B	SY/MD	Ok,M
7	P4467-03	TP-1-VOC	VY019974.D	22 Oct 2024 12:26	vial-B	SY/MD	Ok
8	P4460-02	WB-303-TOP	VY019975.D	22 Oct 2024 12:49	vial-B	SY/MD	Ok
9	P4472-06	BP-F-6-VOC	VY019976.D	22 Oct 2024 13:13	vial-B	SY/MD	Ok
10	P4472-02	BP-F-28-VOC	VY019977.D	22 Oct 2024 13:36	vial-B	SY/MD	Ok
11	VIBLK	VIBLK	VY019978.D	22 Oct 2024 13:59	Clean up	SY/MD	Ok
12	P4471-01	B-180-SB01	VY019979.D	22 Oct 2024 14:23	vial-A	SY/MD	Ok
13	P4471-02	B-180-SB02	VY019980.D	22 Oct 2024 14:46	vial-A Internal standard fail	SY/MD	ReRun
14	P4470-01	CL-01-102124	VY019981.D	22 Oct 2024 15:58	vial-A	SY/MD	Ok
15	P4474-01	TS-2	VY019982.D	22 Oct 2024 16:21	vial-A Internal standard fail	SY/MD	ReRun
16	VSTDCCC050	VSTDCCC050EC	VY019983.D	22 Oct 2024 16:44		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY102324

Review By	Maresh Dadoda	Review On	10/24/2024 2:47:28 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/24/2024 2:50:12 PM
SubDirectory	VY102324	HP Acquire Method	MSVOA_Y HP Processing Method 82y100924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131048
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131049,VP131050 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY019984.D	23 Oct 2024 08:51		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY019985.D	23 Oct 2024 10:58		SY/MD	Ok,M
3	VY1023SBL01	VY1023SBL01	VY019986.D	23 Oct 2024 11:35		SY/MD	Ok
4	VY1023SBS01	VY1023SBS01	VY019987.D	23 Oct 2024 12:20		SY/MD	Ok,M
5	VY1023SBSD01	VY1023SBSD01	VY019988.D	23 Oct 2024 12:43		SY/MD	Ok,M
6	P4487-07	BP-B27-VOC	VY019989.D	23 Oct 2024 13:12	vial-A	SY/MD	Ok
7	P4487-05	BP-F27	VY019990.D	23 Oct 2024 13:35	vial-A	SY/MD	Ok
8	P4487-03	BP-B5-VOC	VY019991.D	23 Oct 2024 13:59	vial-A	SY/MD	Ok
9	P4487-01	BP-B5	VY019992.D	23 Oct 2024 14:22	Internal Standard Fail vial-A	SY/MD	ReRun
10	P4471-02	B-180-SB02	VY019993.D	23 Oct 2024 14:45	vial-B	SY/MD	Ok
11	P4486-01	EO-03-102224	VY019994.D	23 Oct 2024 15:09	Internal Standard Fail vial-A	SY/MD	ReRun
12	P4473-01	TS-1	VY019995.D	23 Oct 2024 15:32	Internal Standard Fail vial-A	SY/MD	ReRun
13	P4474-01RE	TS-2RE	VY019996.D	23 Oct 2024 15:56	Internal Standard Fail vial-B	SY/MD	Confirms
14	P4489-01	RT-2675	VY019997.D	23 Oct 2024 16:19	vial- A Internal Standard Fail; Need MEOH	SY/MD	Dilution
15	VIBLK	VIBLK	VY019998.D	23 Oct 2024 16:43		SY/MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VY019999.D	23 Oct 2024 17:05		SY/MD	Ok,M

M : Manual Integration