

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG NO.: P4471
 Lab File ID: VY019985.D Date Analyzed: 10/23/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	247993	7.71	444115	8.62	376691	11.41
UPPER LIMIT	495986	8.213	888230	9.116	753382	11.914
LOWER LIMIT	123997	7.213	222058	8.116	188346	10.914
EPA SAMPLE NO.						
B-180-SB02	211103	7.71	422929	8.62	336247	11.41
VY1023SBL01	239192	7.71	486232	8.62	427644	11.41
VY1023SBS01	228334	7.71	420481	8.62	359220	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.