

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4471 SAS No.: P4471 SDG NO.: P4471
 Lab File ID: VY019969.D Date Analyzed: 10/22/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:01
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	173474	13.347				
UPPER LIMIT	346948	13.847				
LOWER LIMIT	86737	12.847				
EPA SAMPLE NO.						
B-180-SB01	178248	13.35				
VY1022SBL01	132931	13.35				
VY1022SBS01	160914	13.35				
VY1022SBSD01	163027	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.