

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>Yannuzzi Group, Inc.</u>	SDG No.:	<u>P4473</u>
Contract:	<u>YANN01</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>P4473</u>
		SAS No.:	<u>P4473</u>
		Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	245000	255000	96	216000	294000	10/24/2024	13:43	LB133110
	Antimony	0.022			-50	50	10/24/2024	13:43	LB133110
	Arsenic	4.95			-20	20	10/24/2024	13:43	LB133110
	Barium	3.29	6.0	55	-94	106	10/24/2024	13:43	LB133110
	Beryllium	1.29			-6	6	10/24/2024	13:43	LB133110
	Cadmium	3.37	1.0	337	-5	7	10/24/2024	13:43	LB133110
	Calcium	227000	245000	93	208000	282000	10/24/2024	13:43	LB133110
	Chromium	55.4	52.0	106	42	62	10/24/2024	13:43	LB133110
	Cobalt	2.27			-30	30	10/24/2024	13:43	LB133110
	Copper	8.59	2.0	430	-18	22	10/24/2024	13:43	LB133110
	Iron	101000	101000	100	85600	116500	10/24/2024	13:43	LB133110
	Lead	9.27			-12	12	10/24/2024	13:43	LB133110
	Magnesium	246000	255000	96	216000	294000	10/24/2024	13:43	LB133110
	Manganese	3.89	7.0	56	-13	27	10/24/2024	13:43	LB133110
	Nickel	2.59	2.0	130	-38	42	10/24/2024	13:43	LB133110
	Potassium	-36.7			0	0	10/24/2024	13:43	LB133110
	Selenium	-16.1			-20	20	10/24/2024	13:43	LB133110
	Silver	0.37			-10	10	10/24/2024	13:43	LB133110
	Sodium	4.15			0	0	10/24/2024	13:43	LB133110
	Thallium	3.67			-40	40	10/24/2024	13:43	LB133110
	Vanadium	5.91			-40	40	10/24/2024	13:43	LB133110
	Zinc	5.89			-40	40	10/24/2024	13:43	LB133110
ICSAB01	Aluminum	256000	247000	104	209000	285000	10/24/2024	13:54	LB133110
	Antimony	643	618	104	525	711	10/24/2024	13:54	LB133110
	Arsenic	111	104	107	88.4	120	10/24/2024	13:54	LB133110
	Barium	488	537	91	437	637	10/24/2024	13:54	LB133110
	Beryllium	482	495	97	420	570	10/24/2024	13:54	LB133110
	Cadmium	981	972	101	826	1120	10/24/2024	13:54	LB133110
	Calcium	232000	235000	99	199000	271000	10/24/2024	13:54	LB133110
	Chromium	565	542	104	460	624	10/24/2024	13:54	LB133110
	Cobalt	527	476	111	404	548	10/24/2024	13:54	LB133110
	Copper	518	511	101	434	588	10/24/2024	13:54	LB133110
	Iron	104000	99300	105	84400	114500	10/24/2024	13:54	LB133110
	Lead	57.7	49.0	118	37	61	10/24/2024	13:54	LB133110
	Magnesium	249000	248000	100	210000	286000	10/24/2024	13:54	LB133110
	Manganese	471	507	93	430	584	10/24/2024	13:54	LB133110
	Nickel	1040	954	109	810	1100	10/24/2024	13:54	LB133110
	Potassium	-31.8			0	0	10/24/2024	13:54	LB133110
	Selenium	35.0	46.0	76	26	66	10/24/2024	13:54	LB133110
	Silver	230	201	114	170	232	10/24/2024	13:54	LB133110
	Sodium	-3.02			0	0	10/24/2024	13:54	LB133110
	Thallium	99.3	108	92	68	148	10/24/2024	13:54	LB133110

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Yannuzzi Group, Inc. **SDG No.:** P4473
Contract: YANN01 **Lab Code:** CHEM **Case No.:** P4473 **SAS No.:** P4473
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	460	491	94	417	565	10/24/2024	13:54	LB133110
	Zinc	918	952	96	809	1095	10/24/2024	13:54	LB133110