

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/22/2024

Lab Sample No.: AR1660CCC500 Data File : PP068120.D Time Analyzed: 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.926	5.819	6.019	454.860	500.000	-9.0
Aroclor-1016-2	5.948	5.842	6.042	450.150	500.000	-10.0
Aroclor-1016-3	6.012	5.905	6.105	424.040	500.000	-15.2
Aroclor-1016-4	6.110	6.003	6.203	433.190	500.000	-13.4
Aroclor-1016-5	6.405	6.298	6.498	423.950	500.000	-15.2
Aroclor-1260-1	7.527	7.422	7.622	431.540	500.000	-13.7
Aroclor-1260-2	7.780	7.675	7.875	408.050	500.000	-18.4
Aroclor-1260-3	8.143	8.037	8.237	405.150	500.000	-19.0
Aroclor-1260-4	8.378	8.275	8.475	487.540	500.000	-2.5
Aroclor-1260-5	8.718	8.612	8.812	445.850	500.000	-10.8
Decachlorobiphenyl	10.678	10.569	10.769	43.800	50.000	-12.4
Tetrachloro-m-xylene	4.762	4.655	4.855	53.280	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/22/2024

Lab Sample No.: AR1660CCC500 Data File : PP068120.D Time Analyzed: 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.159	5.061	5.261	449.800	500.000	-10.0
Aroclor-1016-2	5.179	5.081	5.281	481.970	500.000	-3.6
Aroclor-1016-3	5.359	5.262	5.462	482.710	500.000	-3.5
Aroclor-1016-4	5.399	5.301	5.501	444.810	500.000	-11.0
Aroclor-1016-5	5.618	5.520	5.720	460.900	500.000	-7.8
Aroclor-1260-1	6.662	6.564	6.764	464.140	500.000	-7.2
Aroclor-1260-2	6.848	6.750	6.950	451.900	500.000	-9.6
Aroclor-1260-3	7.006	6.908	7.108	465.400	500.000	-6.9
Aroclor-1260-4	7.480	7.383	7.583	460.440	500.000	-7.9
Aroclor-1260-5	7.719	7.622	7.822	479.760	500.000	-4.0
Decachlorobiphenyl	9.212	9.119	9.319	45.900	50.000	-8.2
Tetrachloro-m-xylene	4.051	3.952	4.152	51.090	50.000	2.2