

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/22/2024

Lab Sample No.: AR1660CCC500 Data File : PP068135.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.921	5.819	6.019	476.530	500.000	-4.7
Aroclor-1016-2	5.943	5.842	6.042	476.580	500.000	-4.7
Aroclor-1016-3	6.007	5.905	6.105	445.380	500.000	-10.9
Aroclor-1016-4	6.105	6.003	6.203	435.360	500.000	-12.9
Aroclor-1016-5	6.399	6.298	6.498	426.520	500.000	-14.7
Aroclor-1260-1	7.522	7.422	7.622	469.860	500.000	-6.0
Aroclor-1260-2	7.775	7.675	7.875	457.190	500.000	-8.6
Aroclor-1260-3	8.136	8.037	8.237	431.890	500.000	-13.6
Aroclor-1260-4	8.374	8.275	8.475	443.470	500.000	-11.3
Aroclor-1260-5	8.711	8.612	8.812	453.220	500.000	-9.4
Decachlorobiphenyl	10.666	10.569	10.769	46.360	50.000	-7.3
Tetrachloro-m-xylene	4.757	4.655	4.855	52.520	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/22/2024

Lab Sample No.: AR1660CCC500 Data File : PP068135.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.061	5.261	469.700	500.000	-6.1
Aroclor-1016-2	5.177	5.081	5.281	484.650	500.000	-3.1
Aroclor-1016-3	5.358	5.262	5.462	473.010	500.000	-5.4
Aroclor-1016-4	5.397	5.301	5.501	446.500	500.000	-10.7
Aroclor-1016-5	5.616	5.520	5.720	503.430	500.000	0.7
Aroclor-1260-1	6.659	6.564	6.764	461.890	500.000	-7.6
Aroclor-1260-2	6.846	6.750	6.950	453.780	500.000	-9.2
Aroclor-1260-3	7.003	6.908	7.108	460.980	500.000	-7.8
Aroclor-1260-4	7.477	7.383	7.583	461.370	500.000	-7.7
Aroclor-1260-5	7.715	7.622	7.822	476.330	500.000	-4.7
Decachlorobiphenyl	9.208	9.119	9.319	46.250	50.000	-7.5
Tetrachloro-m-xylene	4.050	3.952	4.152	49.980	50.000	0.0