

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/22/2024

Lab Sample No.: AR1660CCC500 Data File : PP068145.D Time Analyzed: 20:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.921	5.819	6.019	521.080	500.000	4.2
Aroclor-1016-2	5.943	5.842	6.042	525.740	500.000	5.1
Aroclor-1016-3	6.007	5.905	6.105	525.970	500.000	5.2
Aroclor-1016-4	6.105	6.003	6.203	516.400	500.000	3.3
Aroclor-1016-5	6.399	6.298	6.498	489.130	500.000	-2.2
Aroclor-1260-1	7.522	7.422	7.622	483.700	500.000	-3.3
Aroclor-1260-2	7.775	7.675	7.875	478.850	500.000	-4.2
Aroclor-1260-3	8.137	8.037	8.237	484.590	500.000	-3.1
Aroclor-1260-4	8.374	8.275	8.475	477.400	500.000	-4.5
Aroclor-1260-5	8.711	8.612	8.812	478.470	500.000	-4.3
Decachlorobiphenyl	10.665	10.569	10.769	50.750	50.000	1.5
Tetrachloro-m-xylene	4.757	4.655	4.855	53.520	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/22/2024

Lab Sample No.: AR1660CCC500 Data File : PP068145.D Time Analyzed: 20:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.061	5.261	472.880	500.000	-5.4
Aroclor-1016-2	5.177	5.081	5.281	479.700	500.000	-4.1
Aroclor-1016-3	5.358	5.262	5.462	484.480	500.000	-3.1
Aroclor-1016-4	5.397	5.301	5.501	478.080	500.000	-4.4
Aroclor-1016-5	5.616	5.520	5.720	525.350	500.000	5.1
Aroclor-1260-1	6.659	6.564	6.764	470.850	500.000	-5.8
Aroclor-1260-2	6.845	6.750	6.950	468.080	500.000	-6.4
Aroclor-1260-3	7.003	6.908	7.108	466.000	500.000	-6.8
Aroclor-1260-4	7.477	7.383	7.583	462.290	500.000	-7.5
Aroclor-1260-5	7.716	7.622	7.822	465.380	500.000	-6.9
Decachlorobiphenyl	9.207	9.119	9.319	47.030	50.000	-5.9
Tetrachloro-m-xylene	4.050	3.952	4.152	49.440	50.000	-1.1