

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: YANN01
 ProjectID: 86 Davidson Road, Piscataway, NJ
 Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG No.: P4473

Calibration Sequence : FF102124		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	225766981	132804	FF014706.D
850	115534550	135923	FF014707.D
340	48250267	141913	FF014708.D
170	24049317	141467	FF014709.D
85	12677985	149153	FF014710.D
AVG RF : 140252		% RSD : 4.477	AVG RT : 14.9988